



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest

AMCAL CALIBRATION LABS, LLC.

14419 SHAWNEE STREET
MOORPARK, CALIFORNIA 93021, U.S.A.

Calibration Laboratory CL-137

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Expiration Date December 1, 2025

Effective Date February 19, 2025



International Accreditation Service

Issued under the authority of IAS management

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

AMCAL CALIBRATION LABS, LLC.

www.amcallab.com

Contact Name Hillary T. Dinh

Contact Phone +1-818-340-6699

Accredited to ISO/IEC 17025:2017

Effective Date February 19, 2025

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Mechanical			
Pressure Gauges	Up to 300 psi	0.03 psi	EURAMET CG 17 DKD R1
Torque Wrenches / Screwdrivers	Up to 50 lbf ft	0.11 lbf ft	EURAMET CG 14
Precision Balance and Scales	Up to 5 kg 5 kg to 10 kg 10 kg to 25 kg	1 g 2 g 5 g	Direct Method by using standard weights: Class F weights

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

** If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.*

CL-137

AMCAL CALIBRATION LABS, LLC.

Effective Date February 19, 2025

Page 2 of 2

IAS/CL/100-3

