



CERTIFICATE OF ACCREDITATION

This is to attest that

ENERGY DISTRIBUTOR LAB FOR MEASUREMENT AND CALIBRATION

KING ABDUL AZIZ ROAD - AL AMANA DISTRICT - NEAR SAAD BIN MOUAATH MOSQUE
RIYADH, 11372, KINGDOM OF SAUDI ARABIA

Calibration Laboratory CL-182

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date May 5, 2022

Expiration Date November 1, 2024



A handwritten signature in black ink that reads 'Raj Nathan'.

President

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

ENERGY DISTRIBUTOR LAB FOR MEASUREMENT AND CALIBRATION

www.ed.com.sa

Contact Name Monther Waleed

Contact Phone + 966-55-2664411

Accredited to ISO/IEC 17025:2017

Effective Date May 5, 2022

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<i>Dimensional</i>			
External Micrometer ³	0 mm to 300 mm	2.3 µm	Micrometer Checker, Slip Gauge Set Grade 0, Optical Parallel and Optical Flat EDLMC/CM/M/01
Internal Micrometer ³	0 mm to 600 mm	4.3 µm	Micrometer Checker, Slip Gauge Set Grade 0, Optical Parallel and Optical Flat EDLMC/CM/M/01
Caliper ³ (External, Internal)	0 mm to 1000 mm	6.5 µm	Caliper Checker, Slip Gauge Set Grade 0, Optical Parallel EDLMC/CM/M/02
Height gauge ³	0 mm to 1000 mm	9.5 µm	Caliper Checker, Slip Gauge Set Grade 0 EDLMC/CM/M/03
Thickness Gauge ³ (with dial Gauge)	0 mm to 5 mm 0 mm to 25 mm 0 mm to 50 mm	1.1 µm 7.8 µm 8.5 µm	Slip Gauge Set Grade 0 EDLMC/CM/M/04
Feeler Gauge ³	0.01 mm to 2 mm	2.5 µm	Digital Micrometer EDLMC/CM/M/05
Bevel Protractor ³	0° to 360°	10'	Angle Gauges EDLMC/CM/M/21
Dial Indicator ³	0 mm to 25 mm	1.3 µm	Dial Indicator Calibrator EDLMC/CM/M/24

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Profile Projector ³ Linear Angular Magnification	Up to 200 mm Up to 360° 20 X	8 µm 0.02° 0.24 %	Reference Glass Ruler Gauge Block Angle Gauges Digital Caliper EDLMC/CM/M/26
Coating Thickness Masters & Standard Foils ³	0.01 mm to 1 mm	2.5 µm	Digital Micrometer EDLMC/CM/M/06
Coating Thickness Gauge ³	up to 1 mm	2.0 µm	Reference Coating Thickness Material EDLMC/CM/M/27
<i>Mechanical</i>			
Torque Wrench / Torque Device or Meter ³	0.03 N·m to 2 N·m 2 N·m to 10 N·m 10 N·m to 27 N·m 27 N·m to 100 N·m 100 N·m to 2712 N·m	0.6 % 0.4 % 0.4 % 0.3 % 0.6 %	Precision Torque Calibrator Transducers with digital display as per ISO 6789 EDLMC/CM/M/07
Sound Level Meter ³	114 dBA (125 Hz to 2000 Hz)	0.65 dBA	Sound Level Calibrator EDLMC/CM/M/10
Pneumatic Pressure ³ Pressure and Vacuum Gauge (Dial/Digital) and Recorder / Pressure transducer / transmitter / Safety valve	-12 psi to -0.1 psi -0.1 psi to 0.1 psi 0.1 psi to 500 psi 500 psi to 1500 psi -33 mbar to -0.001 mbar -0.001 mbar to 0.001 mbar 0.001 mbar to 33 mbar	0.75 % 0.0075 psi 0.75 % 1.8 % 0.09 % 3.7 µbar 0.09 %	Reference Pressure Calibrator & Reference Pressure Generator EDLMC/CM/M/08
Hydraulic Pressure ³ Pressure and Vacuum Gauge (Dial/Digital) and Recorder / Pressure transducer / transmitter / Safety valve	100 kPa to 7000 kPa 7 MPa to 100 MPa	0.04 % 0.05 %	DWT Electronic Dead Weight Tester EDLMC/CM/M/08
	10 psi to 10000 psi	0.5 %	Reference Pressure Calibrator & Reference Pressure Generator EDLMC/CM/M/08
Portable Vacuum Gauges ³ (Teledyne Hastings)	0.1 mTorr to 100 mTorr (DV5)	18 %	Teledyne Hastings Vacuum Tube Reference Standards DV4, DV5 & DV6 EDLMC/CM/M/34
	1 mTorr to 1 Torr (DV6)	16 %	
	10 mTorr to 10 Torr (DV4)	16 %	
Weight Balance and Scales ³	0 g to 320 g	0.59 mg	E2 Class Weights M1 Class Weights E2 Class Weights E2 Class Weights M1 Class Weights M2 Class Weights M2 Class Weights M2 Class Weights EDLMC/CM/M/09
	0 g to 320 g	3.5 mg	
	0 kg to 24 kg	100 mg	
	0 kg to 110 kg	12 g	
	0 kg to 150 kg	22 g	
	Up to 500 kg	2 kg	
	500 kg to 1000 kg	2.3 kg	
	1000 kg to 9000 kg	3.1 kg	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Force – Tension / Compression (Push/Pull Gauge) ³	0 N to 2000 N 0 lbf to 500 lbf	1.4 N 0.32 lbf	Reference Force Gauge EDLMC/CM/M/29
Force – Tension ³	0 kgf to 1000 kgf 1000 kgf to 5000 kgf 5000 kgf to 10000 kgf 10000 kgf to 20000 kgf 20000 kgf to 30000 kgf 30000 kgf to 40000 kgf 40000 kgf to 50000 kgf	0.60 kgf 1.9 kgf 3.1 kgf 6.2 kgf 9.2 kgf 12 kgf 16 kgf	Comparison with Reference Load Cell As per ISO 376 and OIML R65 (2006) EDLMC/CM/M/18
Force – Compression ³	0 kgf to 1000 kgf 1000 kgf to 5000 kgf 5000 kgf to 10000 kgf 10000 kgf to 20000 kgf 20000 kgf to 30000 kgf 30000 kgf to 40000 kgf 40000 kgf to 50000 kgf	1.1 kgf 1.9 kgf 3.1 kgf 6.2 kgf 9.3 kgf 12 kgf 15 kgf	Comparison with Reference Load Cell As per ISO 376 and OIML R65 (2006) EDLMC/CM/M/18
Load cell-S Type & Dynamometer -Measure ³	100 kgf to 20000 kgf	0.65 %	Reference Load Cell EDLMC/CM/M/14
Durometer ³	30 HA to 90 HA	2 HA	Standard Reference Blocks of Nominal Values 30 HA, 60 HA & 90 HA EDLMC/CM/M/30
Fuel Dispensation By Volume ³ (at Site – in a gasoline station)	Petrol N91: 20 L Petrol N95: 20 L Diesel: 20 L	0.025 L 0.027 L 0.017 L	Volumetric Standards for Petrol (N91 and N95) and Diesel EDLMC/CM/M/45
Anemometer ³	for fixed values of: 492 ft/min 984 ft/min 1969 ft/min 2953 ft/min	4 %	Digital Airflow Meter (By Comparison) EDLMC/CM/M/31
Vibration Meter ³	1 m/s ² to 5 m/s ² 5 m/s ² to 10 m/s ² 10 m/s ² to 20 m/s ²	0.1 m/s ² 0.2 m/s ² 0.6 m/s ²	Vibration Generator EDLMC/CM/M/12
Hydrometer ³	0.7 SG to 1 SG	0.0076 SG	Reference Hydrometer with Baume '10-70 Be' liquid EDLMC/CM/M/40
Liquid Flow Meter ³	2 m ³ /h to 350 m ³ /h	2.2 %	Liquid Flow Meter Comparison Method GE PT878 S/N: GAA1547029 Ultrasonic Flow Meter EDLMC/CM/M/22

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Thermal			
Temperature Controller / Indicator / Recorder with sensor / Glass Thermometer / Digital Thermometer / Thermocouple / Sensor / Temperature Gauge / Transmitter ³	-45 °C to 140 °C 140 °C to 700 °C	0.07 °C 0.43 °C	Metrology Well EDLMC/CM/T/01
Infrared Thermometer / Pyrometer ³	-25 °C to 150 °C 50 °C to 500 °C 500 °C to 980 °C	0.15 °C 0.38 °C 3.1 °C	Fluke 9133 IR Calibrator Fluke 9132 IR Calibrator Omega Black Body Radiation Source EDLMC/CM/T/03
Infrared Calibrator ³	-30 °C to 0 °C 0 °C to 50 °C 50 °C to 250 °C 250 °C to 500 °C	0.32 °C 0.09 °C 0.50 °C 0.54 °C	Comparison with Reference PRT sensor/digital temp indicator as per Euramet cg-13 Version 3(02/2015) EDLMC/CM/T/06
Oven / Incubator / Freezer / Furnaces ³	-50 °C to 1000 °C	1.2 °C	Reference Indicator with Sensor EDLMC/CM/T/02
Dry Block Calibrator ³	-45 °C to 140 °C 140 °C to 700 °C	0.09 °C 0.1 °C	Comparison with Reference PRT sensor/digital temp indicator as per Euramet cg-13 Version 3(02/2015) EDLMC/CM/T/06
Humidity Meter, Sensor with Indicator, Thermo-Hygrometer, Dry Bulb and Wet bulb Thermometer ³	Ambient to 40 °C 20 %RH to 90 %RH	0.27 °C 1.2 %RH	Humidity Chamber, Reference Humidity Monitoring Device EDLMC/CM/T/04
Dew Point Meter ³	- 50 °C to 20 °C	0.2 °C	Reference Dew Point Meter EDLMC/CM/T/05
RTD Pt 100 (Probe Calibration) ³	-45 °C to 700 °C	0.09 °C	Reference Multimeter Temperature / Pressure Calibrator and Digital Thermometer Readout Metrology Well EDLMC/CM/T/01
Electrical – DC/LF			
DC Voltage – Source ^{3,4}	0.1 µV to 300 mV 300 mV to 3 V 3 V to 30 V 30 V to 300 V 300 V to 1000 V	5 µV/V + 1 µV 5 µV/V + 10 µV 7 µV/V + 100 µV 17 µV/V + 1 mV 15 µV/V + 5 mV	Multi-Function Calibrator by Direct Method EDLMC/CM/E/01

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Current - Source ^{3,4}	1 nA to 300 µA 300 µA to 3 mA 3 mA to 30 mA 30 mA to 300 mA 300 mA to 3 A 3 A to 20 A 20 A to 1000 A	260 µA/A + 2 nA 690 µA/A + 20 nA 270 µA/A + 0.2 µA 160 µA/A + 2 µA 460 µA/A + 20 µA 3 mA/A + 200 µA 3.8 mA/A	Multi-Function Calibrator by Direct Method EDLMC/CM/E/03 Multi-Function Calibrator by Direct Method with 2/10/50 Turn Current Coil
AC Voltage - Source ^{3,4}	1 mV to 33 mV (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz) (at 500 kHz) 33 mV to 330 mV (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz) (at 500 kHz) 330 mV to 3.3 V (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 500 kHz) (at 500 kHz) 3.3 mV to 33 V (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (at 100 kHz)	0.18 % 0.17 % 0.16 % 0.16 % 0.17 % 0.18 % 0.19 % 0.25 % 0.093 % 0.093 % 0.093 % 0.093 % 0.093 % 0.095 % 0.098 % 0.11 % 0.004 % 0.002 % 0.002 % 0.002 % 0.003 % 0.005 % 0.012 % 0.023 % 0.012 % 0.005 % 0.005 % 0.008 % 0.015 % 0.039 % 0.048 %	Multi-Function Calibrator by Direct Method EDLMC/CM/E/02

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)		
AC Voltage - Source ^{3,4} continued	33 V to 330 V (45 Hz to 1 kHz)	0.006 %	Multi-Function Calibrator by Direct Method		
	(1 kHz to 10 kHz)	0.005 %			
	(10 kHz to 20 kHz)	0.005 %			
	(20 kHz to 50 kHz)	0.006 %			
	33 V to 330 V (50 kHz to 100 kHz)	0.008 %			
	(at 100 kHz)	0.02 %			
	330 V to 1020 V (45 Hz to 1 kHz)	0.009 %			
	(1 kHz to 5 kHz)	0.008 %			
	(5 kHz to 10 kHz)	0.008 %			
	(at 10 kHz)	0.01 %			
	AC Current - Source ^{3,4}	29 µA to 329 µA (10 Hz to 20 Hz)		0.11 %	Multi-Function Calibrator by Direct Method EDLMC/CM/E/04
		(20 Hz to 45 Hz)		0.11 %	
(45 Hz to 1 kHz)		0.09 %			
(1 kHz to 5 kHz)		0.09 %			
(5 kHz to 10 kHz)		0.10 %			
(10 kHz to 30 kHz)		0.13 %			
(at 30 kHz)		0.19 %			
330 µA to 3.3 mA (10 Hz to 20 Hz)		0.08 %			
(20 Hz to 45 Hz)		0.08 %			
(45 Hz to 1 kHz)		0.04 %			
(1 kHz to 5 kHz)		0.04 %			
(5 kHz to 10 kHz)		0.05 %			
(10 kHz to 30 kHz)		0.12 %			
(at 30 kHz)		0.18 %			
3.3 mA to 33 mA (10 Hz to 20 Hz)		0.05 %			
(20 Hz to 45 Hz)		0.04 %			
(45 Hz to 1 kHz)		0.03 %			
(1 kHz to 5 kHz)		0.02 %			
(5 kHz to 10 kHz)		0.04 %			
(10 kHz to 30 kHz)		0.09 %			
(at 30 kHz)		0.30 %			
33 mA to 330 mA (10 Hz to 20 Hz)		0.05 %			
(20 Hz to 45 Hz)		0.04 %			
(45 Hz to 1 kHz)		0.03 %			
(1 kHz to 5 kHz)	0.03 %				

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Current - Source ^{3,4} continued	33 mA to 330 mA (5 kHz to 10 kHz) (10 kHz to 30 kHz) (at 30 kHz)	0.04 % 0.09 % 0.40 %	Multi-Function Calibrator by Direct Method
	330 mA to 1.1 A (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (at 10 kHz)	0.05 % 0.03 % 0.03 % 0.04 % 0.08 %	
	1.1 A to 3 A (10 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (at 10 kHz)	0.03 % 0.03 % 0.03 % 0.04 % 0.10 %	
	3 A to 20 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz) (at 5 kHz)	0.044 % 0.04 % 0.04 % 0.083 %	
	20 A to 1000 A (45 Hz to 1 kHz)	0.38 %	
Frequency - Source ^{3,4}	0.3 Hz to 1200 kHz 1.2 MHz to 2 MHz	0.002 % 0.004 %	Multi-Function Calibrator by Direct Method EDLMC/CM/E/09
Capacitance – Source ^{3,4} (100 Hz to 1 kHz)	220 pF to 400 pF 0.6 nF to 330 nF 0.33 µF to 330 µF 0.33 mF to 110 mF	5 % 0.006 % 0.011 % 0.064 %	Multi-Function Calibrator by Direct Method EDLMC/CM/E/06
Energy Meter ³ (50 Hz / 60 Hz)	1 Wh to 10 kWh 10 W to 10 kW 1 VA to 10 kVA 20 V to 300 V 500 mA to 100 A	0.75 Wh 0.17 % 0.17 % 0.011 % 0.054 %	Three Phase Reference Meter and Portable Power Source EDLMC/CM/E/35

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Temperature Simulation – Measure/Source ³ (Temperature Indicator / Controller / Recorder / Test Kit / Universal Calibrator)			Multi-Function Calibrator or Process Calibrator
RTD – Pt 100	-200 °C to 800 °C	0.09 °C	EDLMC/CM/E/07
Thermocouples			Multi-Function Calibrator or Process Calibrator
Type – B	600 °C to 1820 °C	0.50 °C	EDLMC/CM/E/08
Type – E	-250 °C to 1000 °C	0.57 °C	
Type – J	-210 °C to 1200 °C	0.31 °C	
Type – K	-200 °C to 1372 °C	0.48 °C	
Type – L	-200 °C to 900 °C	0.43 °C	
Type – N	-200 °C to 1300 °C	0.46 °C	
Type – R	1 °C to 1767 °C	0.66 °C	
Type – S	1 °C to 1767 °C	0.56 °C	
Type – T	-250 °C to 400 °C	0.73 °C	
Type – U	-200 °C to 600 °C	0.65 °C	
Oil Tester ³	1 kV to 100 kV	0.6 kV	Oil Tester Reference Meter EDLMC/CM/E/28
AC Power – Source ^{3,4} (1-Phase, at UPF)	100 mW to 20 kW (60 Hz, 1 V to 1000 V) (500 Hz, 1 V to 1000 V) (1 kHz, 1 V to 1000 V)	0.9 % 0.9 % 0.9 %	Multifunction Calibrator 2/10/50 Turn Current Coil EDLMC/CM/E/30
AC Power – Measure ^{3,5} (1-Phase, at UPF)	500 mW to 5 kW (60 Hz, 120 V/240 V)	0.03 %	Reference Single Phase Meter EDLMC/CM/E/38
DC Power – Source ^{3,4}	100 mW to 1 W (1 V / 100 mA to 1 A)	0.06 %	Multifunction Calibrator EDLMC/CM/E/29
	1 W to 20 kW (1 V to 1 kV / 1 A to 20 A)	1.1 %	
DC Voltage – Measure ^{3,5}	1 mV to 10 mV 10 mV to 100 mV 100 mV to 100 V 100 V to 1000 V	0.0081 % 0.00033 % 0.00033 % 0.00027 %	Reference Digital Multimeter EDLMC/CM/E/10
	1000 V to 200 kV 200 kV to 300 kV	0.19 % 0.39 %	High Voltage Reference Meter EDLMC/CM/E/10
AC Voltage – Measure ^{3,5}	1 mV to 200 mV (1 Hz to 20 Hz) (20 Hz to 3 kHz) (3 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 60 kHz) (60 kHz to 100 kHz)	0.0019 % 0.0019 % 0.0019 % 0.0019 % 0.0019 % 0.0021 %	Reference Digital Multimeter EDLMC/CM/E/11

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage – Measure ^{3,5} continued	200 mV to 2 V		Reference Digital Multimeter
	(1 Hz to 20 Hz)	0.0018 %	
	(20 Hz to 3 kHz)	0.0018 %	
	(3 kHz to 10 kHz)	0.0018 %	
	(10 kHz to 30 kHz)	0.0018 %	
	(30 kHz to 60 kHz)	0.0018 %	
	(60 kHz to 100 kHz)	0.0018 %	
	2 V to 20 V		
	(1 Hz to 20 Hz)	0.0022 %	
	(20 Hz to 55 Hz)	0.0022 %	
	(55 Hz to 1 kHz)	0.0022 %	
	(1 kHz to 3 kHz)	0.0022 %	
	(3 kHz to 10 kHz)	0.0022 %	
	(10 kHz to 30 kHz)	0.0021 %	
	(30 kHz to 100 kHz)	0.0035 %	
(100 kHz to 500 kHz)	0.085 %		
(500 kHz to 1 MHz)	0.085 %		
20 V to 200 V			
(1 Hz to 20 Hz)	0.0021 %		
(20 Hz to 55 Hz)	0.0021 %		
(55 Hz to 1 kHz)	0.0021 %		
(1 kHz to 3 kHz)	0.0021 %		
(3 kHz to 10 kHz)	0.0021 %		
(10 kHz to 30 kHz)	0.007 %		
(30 kHz to 60 kHz)	0.020 %		
(60 kHz to 100 kHz)	0.020 %		
200 V to 1000 V			
(1 Hz to 55 Hz)	0.0021 %		
(55 Hz to 1 kHz)	0.0021 %		
(1 kHz to 3 kHz)	0.0021 %		
(3 kHz to 10 kHz)	0.0021 %		
1 kV to 200 kV		0.11 %	High Voltage Reference Meter EDLMC/CM/E/11
(1 Hz to 1 MHz)			
200 kV to 300 kV		0.47 %	
DC Current – Measure ^{3,5}	1 nA to 10 mA	6 µA/A	Reference Digital Multimeter EDLMC/CM/E/12
	10 mA to 100 mA	13 µA/A	
	100 mA to 1 A	57 µA/A	
	1 A to 20 A	3 µA/A	
	20 A to 600 A	120 µA/A	Current Shunts and Clamp Meter EDLMC/CM/E/12
	600 A to 1000 A	3.2 mA/A	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Current - Measure ^{3,5} (1 Hz to 100 kHz)	2 μ A to 20 A	0.0034 %	Reference Digital Multimeter EDLMC/CM/E/13
	20 A to 600 A	5.8 mA/A	Reference Multimeter with Reference Current Shunts EDLMC/CM/E/13
	600 A to 2000 A	23 mA/A	Reference Clamp Meter EDLMC/CM/E/13
	2000 A to 6000 A	0.30 %	Reference Flexible Current Probe with Indicator EDLMC/CM/E/13
DC Resistance – Source ^{3,4}	0.1 Ω to 100 k Ω	1.1 %	Multifunction Calibrator & 4 Wire EDLMC/CM/E/05
	100 k Ω to 1 G Ω	0.23 %	Multifunction Calibrator & 2 Wire EDLMC/CM/E/05
	10 M Ω to 10 G Ω	0.18 %	Insulation Tester Calibrator EDLMC/CM/E/40
DC Resistance - Measure ^{3,5}	1 $\mu\Omega$ to 20 G Ω	15 $\mu\Omega/\Omega$	Reference Digital Multimeter EDLMC/CM/E/14
Frequency – Measure ^{3,5}	0 Hz to 1 MHz	0.006 %	Reference Digital Multimeter EDLMC/CM/E/18
	1 MHz to 2 MHz	0.006 %	Universal Counter EDLMC/CM/E/18
Capacitance – Measure ^{3,5} (at 1 kHz)	100 pF to 10 μ F	0.8 %	Reference LCR Meter
	10 μ F to 100 mF	2.8 %	Reference Digital Multimeter EDLMC/CM/E/15
Inductance – Measure ^{3,5} (at 1 kHz)	100 μ H to 1 mH	0.8 %	Reference LCR Meter
	1 mH to 10 H	5.7 %	EDLMC/CM/E/20
Transformer Turn Ratio Meter ³	0.8 R to 2000 R (at 8 V, 40 V, 80 V)	0.25 R	3 Phase TTR Calibration Standard EDLMC/CM/E/21
Current Transformer Analyzer ³	CT RATIOS: 1, 5, 2000	0.02 %	Reference Current Transformer EDLMC/CM/E/22
<i>Time and Frequency</i>			
Stop Watch / Timer ³	10 ms to 1 s	0.17 ms	Reference Electronic Timer EDLMC/CM/E/33
	1 s to 900 s	2 ms	Reference Stop Watch

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Tachometer (Contact / Non-Contact) ³	100 rpm to 12000 rpm	1.3 rpm	Tachometer Calibrator (Contact Type) EDLMC/CM/E/11
	100 rpm to 99000 rpm	2.4 rpm	Tachometer Calibrator (Non-Contact Type or Strobe) EDLMC/CM/M/11
Optical Radiation			
Lux/ Light/ Illumination Meter ³	1 lux to 5 lux	5.5 %	Labsphere Luminance Source USS-800C-75R EDLMC/CM/M/13
	5 lux to 100 lux	1.6 %	
	100 lux to 200 lux	1.6 %	
	200 lux to 1000 lux	1.7 %	
	1000 lux to 2000 lux	2.1 %	
	2000 lux to 10000 lux	1.3 %	
	10000 lux to 24000 lux	2.6 %	
Ultra Violet (UV) Meter ³	0.01 mW/cm ² to 1999.9 mW/cm ²	5.8 %	Bachur & Associates UV Source and G & R Labs Reference Meter EDLMC/CM/M/17
Optical Power Meter ³	-60 dBm to 0 dBm (For wavelength = 1310 nm & 1550 nm)	0.1 dBm	Reference Optical Power Meter and Optical Attenuator EDLMC/CM/E/25
Optical Time Domain Reflectometer (OTDR) ³	(At 1310 nm, 1550 nm & 1650 nm) Up to 3 km 3 km to 25 km	0.36 m/km 0.14 m/km	Reference Optical Time Domain Reflectometer (OTDR) EDLMC/CM/E/27
Laser Source (Wavelength) ³	1300 nm 1310 nm 1550 nm	1 nm 1 nm 1 nm	Optical Spectrum Analyzer EDLMC/CM/E/26
Chemical/Gas			
Gas Detector Analyzer ³ Hydrogen Sulfide Carbon Monoxide Methane Oxygen Hydrogen	0 ppm to 200 ppm 0 ppm to 1000 ppm 10 % to 100 % 5 % to 30 % 0 ppm to 1000 ppm	3.6 parts per 10 ⁶ 2.2 parts per 10 ⁶ 3.6 % 2.2 % 9.9 parts per 10 ⁶	Reference Gas Mixture of H ₂ S, CO, CH ₄ , O ₂ and H ₂ EDLMC/CM/M/33
SF ₆ Gas Detector / Analyzer ³	Dew Point - 50 °C to 20 °C	0.2 °C	Dew Point Mirror Standard EDLMC/CM/T/05
	Sulfur dioxide (SO ₂) 9 ppm to 500 ppm	3 % (of reported parts per 10 ⁶)	Reference Gas EDLMC/CM/M/33
	Sulfur hexafluoride 80 % to 100 %	0.15 % (of reported percentage)	

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
pH Meter ³	4.00 pH 7.00 pH 10.00 pH	0.03 pH 0.03 pH 0.03 pH	4.00, 7.00 and 10.00 pH Certified Buffer Solutions EDLMC/CM/M/15
Biomedical			
Defibrillator (For Energy Calibration) ³	1 J to 10 J 10 J to 50 J 50 J to 100 J 100 J to 170 J 170 J to 300 J	0.8 J 1.5 J 2.8 J 5.4 J 8.1 J	Electrical Safety Tester & Defibrillator Analyzer EDLMC/CM/B/10

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Calibration for this parameter/calibration item is available at the laboratory or at customer site.

⁴Capability is suitable for the calibration of measuring devices in the stated ranges.

⁵Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

SG = specific gravity
UPF = unity power factor
R = turn ratio
ppm = parts per million
HA = Shore Hardness Scale A