



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

QATAR AIRWAYS

HAMAD INTERNATIONAL AIRPORT
DOHA 22550, QATAR

Calibration Laboratory CL-205

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date July 10, 2024

Expiration Date May 1, 2025



A handwritten signature in black ink, reading 'Raj Nathan'.

President

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

QATAR AIRWAYS

www.qatarairways.com

Contact Name Mr. Simon Christopher Trillo

Contact Phone + 974-4010-9086

Accredited to ISO/IEC 17025:2017

Effective Date July 10, 2024

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|--|---|--|--|
| <i>Dimensional</i> | | | |
| Micrometer | 0 mm to 25 mm 25 mm to 50 mm 50 mm to 100 mm 75 mm to 100 mm 100 mm to 200 mm | 1.0 µm 1.3 µm 1.5 µm 1.9 µm 2.7 µm | Using Grade 0 Gage Blocks By Direct method |
| Length Measuring Machine | 0.5 mm to 50 mm 50 mm to 200 mm | 0.16 µm 0.25 µm | Using Grade 0 Gage Blocks by Direct method |
| Vernier Caliper (External Jaws) | 0 mm to 300 mm | 6.2 µm | Using Grade 0 Gage Block by Direct Method |
| Dial Gauges for Linear Measurement (Plunger) | 0 mm to 25 mm | 1.3 µm | Using Length Measuring Machine by Direct Method |
| Dial Test Indicators (Lever) | 0 mm to 2 mm | 0.5 µm | Using Length Measuring Machine by Direct Method |
| Depth Gauge | Up to 200 mm | 6 µm | Using Grade 0 Gage Block by Direct method |
| Setting Rod, Pin Gage, Plug Gage | 0.1 mm to 100 mm 100 mm to 200 mm | 0.7 µm 1.5 µm | Using Length Measuring Machine by Direct Method |
| Setting Ring | 6 mm to 20 mm 20 mm to 90 mm | 0.2 µm 0.3 µm | Using Length Measuring Machine by Direct Method |
| 3 Point (Bore) Micrometers | 6 mm to 40 mm 40 mm to 95 mm | 1.1 µm 2.2 µm | Using Setting Rings by Direct method |
| Inclinometer | 0° to 45° | 36" | Using Gage blocks and Sine bar by Direct method |

* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|--|---|--|---|
| Mechanical | | | |
| Pressure Pneumatic ⁵ (Pressure Gauge / Pressure Transmitter) | 1 kPa to 5 kPa | 2.5 % | Using Reference Gauge by Comparison Method |
| | 5 kPa to 350 kPa | 0.1 % | |
| | 350 kPa to 700 kPa | 0.05 % | |
| | 700 kPa to 1750 kPa | 0.06 % | |
| | 1750 kPa to 3500 kPa | 0.02 % | |
| | 3500 kPa to 5.6 MPa 5.6 MPa to 14 MPa | 0.06 % 0.04 % | |
| Pressure Hydraulic ⁵ (Pressure Gauge / Pressure Transmitter) | 2 MPa to 4 MPa | 0.14 % | Using Electronic Deadweight Tester by Direct Method |
| | 4 MPa to 7 MPa | 0.08 % | |
| | 7 MPa to 12 MPa | 0.06 % | |
| | 12 MPa to 20 MPa | 0.03 % | Using Reference Gauge by Comparison Method |
| | 20 MPa to 140 MPa | 0.05 % | |
| Vacuum Gauge ⁵ | 0.1 MPa to 5 MPa | 0.4 % | Using Reference Gauge by Comparison Method |
| | 5 MPa to 10 MPa | 0.2 % | |
| | 10 MPa to 30 MPa | 0.1 % | |
| | 30 MPa to 70 MPa | 0.05 % | |
| Torque (CW and CCW) | 0.1 kPa to 20 kPa | 0.3 % | Using Reference Gauge by Comparison Method |
| | 20 KPa to 30 KPa | 0.2 % | |
| | 30 KPa to 80 KPa | 0.1 % | |
| Torque (CW and CCW) | 0.2 N.m to 1 N.m | 0.76 % | Method: QTR/CAL/054/014 (ISO 6789:2017) Torque Transducers |
| | 1 N.m to 110 N.m | 0.6 % | |
| | 110 N.m to 1100 N.m | 0.5 % | |
| Force-Compression (Aircraft Scales) | 26 kN to 267 kN | 0.06 % | Using Reference load cell and Indicator & Force Generation Machine by Comparison Method |
| Hanging Balances Spring Balances Push-Pull Gauges | 10 N to 1000 N | 0.06 % | Using Suspension weights by Direct Method |
| Weighing scales ⁵ | 10 g to 200 g 201 g to 500 g 501 g to 1 kg 1 kg to 10 kg 10 kg to 60 kg | 5 mg + 0.58R 10 mg + 0.58R 68 mg + 0.8R 0.14 g + 0.8R 0.76 g + 0.8R where R is resolution of unit under calibration | Using F1 Class Weights by ABBA Method |
| Sound Level Meter @ 1 kHz | 94 dB and 114 dB | 0.17 dB | Using Sound level Calibrator by Direct Method |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|---|--|---|--|
| Thermal | | | |
| Temperature Sensors/ Thermometer with or without Indicators ⁵ | -25 °C to 100 °C 100 °C to 400 °C 400 °C to 600 °C | 0.08 °C 0.12 °C 0.58 °C | Using Reference PRT with readout & Drywell Temperature Calibrator by Comparison Method |
| Pyrometers / Non-contact Thermometers ⁵ | 35 °C to 50 °C 50 °C to 100 °C 100 °C to 200 °C 200 °C to 300 °C 300 °C to 400 °C | 0.4 °C 0.7 °C 1.0 °C 1.6 °C 2.1 °C | Using Infrared Calibrator by Direct Method |
| Temperature Sourcing Devices and Enclosures ⁵ (Single Sensor Method) | -25 °C to 200 °C 200 °C to 400 °C 400 °C to 500 °C 500 °C to 600 °C | 0.02 °C 0.03 °C 0.04 °C 0.78 °C | Using PRT/ N Type Thermocouple and Readout By Direct Method |
| RH Meter/ Sensors | 20 %RH to 30 %RH (@ 40 °C) 30 %RH to 75 %RH (@ 25 °C) | 1.4 %RH 1.8 %RH | Using Reference Sensor by Comparison Method |
| Electrical – DC/LF | | | |
| DC Voltage Generate ^{3,5} | 0 mV to 3.29999 mV 3.29999 mV to 329.9999 mV 329.9999 mV to 3.299999 V 3.299999 V to 32.99999 V 32.99999 V to 329.9999 V 329.9999 V to 1000.000 V | 0.0016 % + 0.93 µV 0.0016 % + 3.2 µV 0.00085 % + 16 µV 0.00093 % + 190 µV 0.0014 % + 2.6 mV 0.0014 % + 8.1 mV | Using Electrical Multifunction Calibrator by Direct Method |
| DC Resistance Generate ^{3,5} | 0 Ω to 10.9999 Ω 10.9999 Ω to 32.9999 Ω 32.9999 Ω to 109.9999 Ω 109.9999 Ω to 1.099999 kΩ 1.099999 kΩ to 3.299999 kΩ 3.299999 kΩ to 10.99999 kΩ 10.99999 kΩ to 32.99999 kΩ 32.99999 kΩ to 109.9999 kΩ 109.9999 kΩ to 329.9999 kΩ 329.9999 kΩ to 1.099999 MΩ 1.099999 MΩ to 3.299999 MΩ 3.299999 MΩ to 10.99999 MΩ 10.99999 MΩ to 32.99999 MΩ 32.99999 MΩ to 109.9999 MΩ 109.9999 MΩ to 329.9999 MΩ 329.9999 MΩ to 1000 MΩ | 0.0031 % + 0.98 mΩ 0.0023 % + 1.5 mΩ 0.0022 % + 1.9 mΩ 0.0022 % + 8.3 mΩ 0.0022 % + 35 mΩ 0.0022 % + 84 mΩ 0.0022 % + 0.40 Ω 0.0022 % + 0.84 Ω 0.0025 % + 5.1 Ω 0.0025 % + 50 Ω 0.0047 % + 0.28 kΩ 0.010 % + 1.0 kΩ 0.019 % + 15 kΩ 0.039 % + 80 kΩ 0.23 % + 0.14 MΩ 1.2 % + 0.71 MΩ | Using Electrical Multifunction Calibrator by Direct Method |
| DC Current Generate ^{3,5} | 0 µA to 329.999 µA 329.999 µA to 3.29999 mA 3.29999 mA to 32.9999 mA | 0.012 % + 20 nA 0.0078 % + 86 nA 0.0078 % + 0.78 µA | Using Electrical Multifunction Calibrator by Direct Method |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|---|--|--|---|
| DC Current Generate ^{3,5} (continued) | 32.9999 mA to 329.999 mA 329.999 mA to 1.09999 A 1.09999 A to 2.99999 A 2.99999 A to 10.0000 A 10.0000 A to 20.0000 A | 0.0078 % + 7.9 µA 0.016 % + 59 µA 0.029 % + 0.24 mA 0.039 % + 0.89 mA 0.078 % + 0.58 mA | Using Electrical Multifunction Calibrator by Direct Method |
| DC Current Generate ^{3,5} | 10 A to 100 A 100 A to 500 A 500 A to 1000 A | 0.40 % + 0.07 A 0.44 % + 0.099 A 0.40 % + 0.14 A | Using Electrical Multifunction Calibrator & 50 Turn Current Coil by Direct Method |
| Capacitance Generate ^{3,5} | 220.0 pF to 399.9 pF (10 Hz to 10 kHz) 399.9 pF to 1.0999 nF (10 Hz to 10 kHz) 1.0999 nF to 3.2999 nF (10 Hz to 3 kHz) 3.2999 nF to 10.9999 nF (10 Hz to 1 kHz) 10.9999 nF to 32.9999 nF (10 Hz to 1 kHz) 32.9999 nF to 109.999 nF (10 Hz to 1 kHz) 109.999 nF to 329.999 nF (10 Hz to 1 kHz) 329.99 nF to 1.09999 µF (10 Hz to 600 Hz) 1.09999 µF to 3.29999 µF (10 Hz to 300 Hz) 3.29999 µF to 10.9999 µF (10 Hz to 150 Hz) 10.9999 µF to 32.9999 µF (10 Hz to 120 Hz) 32.9999 µF to 109.999 µF (10 Hz to 80 Hz) 109.999 µF to 329.999 µF (0 Hz to 50 Hz) | 0.39 % + 8.0 pF 0.39 % + 7.9 pF 0.39 % + 11 pF 0.19 % + 14 pF 0.19 % + 63 pF 0.19 % + 99 pF 0.19 % + 0.63 nF 0.19 % + 1.2 nF 0.19 % + 6.7 nF 0.19 % + 13 nF 0.31 % + 75 nF 0.35 % + 97 nF 0.35 % + 0.63 µF | Using Electrical Multifunction Calibrator by Direct Method |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|--|--|--------------------------------|--|
| Capacitance Generate ^{3,5} (continued) | 329.999 µF to 1.09999 mF (0 Hz to 20 Hz) | 0.35 % + 1 nF | Using Electrical Multifunction Calibrator by Direct Method |
| | 1.09999 mF to 3.29999 mF (0 Hz to 6 Hz) | 0.35 % + 6.3 µF | |
| | 3.29999 mF to 10.9999 mF (0 Hz to 2 Hz) | 0.35 % + 10 µF | |
| | 10.9999 mF to 32.9999 mF (0 Hz to 0.6 Hz) | 0.58 % + 63 µF | |
| | 32.9999 mF to 100.00 mF (0 Hz to 0.2 Hz) | 0.85 % + 0.17 mF | |
| AC Voltage Generate ^{3,5} | 10.0 mV to 32.999 mV (45 Hz to 1 kHz) | 0.062 % + 5 µV | Using Electrical Multifunction Calibrator by Direct Method |
| | (1 kHz to 10 kHz) | 0.016 % + 5 µV | |
| | (10 kHz to 20 kHz) | 0.62 % + 5 µV | |
| | (20 kHz to 50 kHz) | 0.27 % + 10 µV | |
| | (50 kHz to 100 kHz) | 0.62 % + 39 µV | |
| | 32.999 mV to 329.999 mV (45 Hz to 1 kHz) | 0.011 % + 11 µV | |
| | (1 kHz to 10 kHz) | 0.012 % + 11 µV | |
| | (10 kHz to 20 kHz) | 0.027 % + 11 µV | |
| | (20 kHz to 50 kHz) | 0.062 % + 26 µV | |
| | (50 kHz to 100 kHz) | 0.16 % + 55 µV | |
| | 329.999 mV to 3.29999 V (45 Hz to 1 kHz) | 0.012 % + 0.10 mV | |
| | (1 kHz to 10 kHz) | 0.015 % + 0.10 mV | |
| | (10 kHz to 20 kHz) | 0.023 % + 0.10 mV | |
| | (20 kHz to 50 kHz) | 0.054 % + 0.13 mV | |
| | (50 kHz to 100 kHz) | 0.19 % + 0.48 mV | |
| | 3.29999 V to 32.9999 V (45 Hz to 1 kHz) | 0.012 % + 0.10 mV | |
| | (1 kHz to 10 kHz) | 0.019 % + 0.98 mV | |
| | (10 kHz to 20 kHz) | 0.027 % + 1.0 mV | |
| | (20 kHz to 50 kHz) | 0.070 % + 1.5 mV | |
| | (50 kHz to 100 kHz) | 0.070 % + 1.5 mV | |
| 32.9999 V to 329.999 V (45 Hz to 1 kHz) | 0.016 % + 10 mV | | |
| (1 kHz to 10 kHz) | 0.019 % + 10 mV | | |
| (10 kHz to 20 kHz) | 0.023 % + 10 mV | | |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|---|---|------------------------------------|---|
| AC Voltage Generate ^{3,5} (continued) | 32.9999 V to 329.999 V (20 kHz to 50 kHz) (50 kHz to 100 kHz) | 0.16 % + 41 mV 0.16 % + 41 mV | Using Electrical Multifunction Calibrator by Direct Method |
| | 329.999 V to 1000.00 V (45 Hz to 1 kHz) (1 kHz to 10 kHz) | 0.023 % + 30 mV 0.023 % + 30 mV | |
| AC Current Generate ^{3,5} | 100 µA to 329.99 µA (45 Hz to 1 kHz) | 0.078 % + 0.16 µA | Using Electrical Multifunction Calibrator by Direct Method |
| | 329.99 µA to 3.29999 mA (45 Hz to 1 kHz) | 0.031 % + 1.8 µA | |
| | 3.29999 mA to 32.9999 mA (45 Hz to 1 kHz) | 0.031 % + 18 µA | |
| | 32.9999 mA to 100.000 mA (45 Hz to 1 kHz) | 0.039 % + 0.018 mA | |
| | 100.000 mA to 1.00000 A (45 Hz to 1 kHz) | 0.039 % + 0.15 mA | |
| | 1.00000 A to 2.99999 A (45 Hz to 1 kHz) | 0.047 % + 3.0 mA | |
| | 2.99999 A to 10.0 A (45 Hz to 1 kHz) | 0.078 % + 3.4 mA | |
| 10.0 A to 20.0 A (45 Hz to 1 kHz) | 0.12 % + 7.2 mA | | |
| AC Current Generate ^{3,5} | 1 A to 10 A (45 Hz to 100 Hz) (100 Hz to 440 Hz) | 0.67 % + 58 mA 1.0 % + 58 mA | Using Electrical Multifunction Calibrator & 50 Turn Current Coil by Direct Method |
| | 10 A to 100 A (45 Hz to 100 Hz) (100 Hz to 440 Hz) | 0.82 % + 58 mA 1.2 % + 58 mA | |
| | 100 A to 200 A (45 Hz to 100 Hz) (100 Hz to 440 Hz) | 0.78 % + 58 mA 1.4 % + 58 mA | |
| | 200 A to 300 A (45 Hz to 100 Hz) (100 Hz to 440 Hz) | 0.67 % + 0.16 A 1.3 % + 0.16 A | |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|---|--|---|---|
| AC Current Generate ^{3,5} (continued) | 300 A to 400 A (45 Hz to 100 Hz) (100 Hz to 440 Hz) | 0.61 % + 0.16 A 1.2 % + 0.16 A | Using Electrical Multifunction Calibrator & 50 Turn Current Coil by Direct Method |
| | 400 A to 500 A (45 Hz to 100 Hz) | 0.92 % + 0.16 A | |
| | 500 A to 1000 A (45 Hz to 100 Hz) | 0.92 % + 0.31 A | |
| Temperature Simulation ⁵ Thermocouple | | | Using Electrical Multifunction Calibrator by Direct Method |
| Type - B | 600 °C to 1750 °C | 0.67 °C | |
| Type - E | -250 °C to 200 °C | 0.39 °C | |
| Type - J | -210 °C to 1190 °C | 0.21 °C | |
| Type - K | -200 °C to 1370 °C | 0.31 °C | |
| Type - N | -200 °C to 600 °C | 0.31 °C | |
| Type - R | 0 °C to 500 °C | 0.45 °C | |
| Type - S | 600 °C to 1200 °C | 0.44 °C | |
| Type - T | -250 °C to 200 °C | 0.49 °C | |
| Insulation DC Resistance Generate ^{3,5} | 100 KΩ to 10 GΩ (Discrete steps) | 1.3 % | Using Standard Resistors by Direct Method |
| DC Resistance Generate ^{3,5} | 50 μΩ to 200 μΩ 200 μΩ to 1 mΩ 1 mΩ to 2 mΩ 2 mΩ to 5 mΩ 5 mΩ to 10 mΩ 10 mΩ to 15 mΩ 15 mΩ to 20 mΩ 20 mΩ to 50 mΩ 50 mΩ to 100 mΩ 100 mΩ to 150 mΩ 150 mΩ to 200 mΩ 200 mΩ to 500 mΩ 500 mΩ to 1000 mΩ 1 Ω to 1.5 Ω 1.5 Ω to 2 Ω | 0.12 μΩ 0.7 μΩ 0.85 μΩ 1.3 μΩ 2.1 μΩ 3.0 μΩ 4.0 μΩ 9.7 μΩ 20 μΩ 30 μΩ 39 μΩ 96 μΩ 0.19 mΩ 0.29 mΩ 0.40 mΩ | Using Standard Resistors and DMM by Comparison |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|--|---|--|---|
| AC Power (Active) Generate ^{3,5} | 10 W to 1000 W (45 Hz to 1 kHz, Power factor = 1) | 0.08 % + 0.58R | Using Electrical Multifunction Calibrator by Direct Method |
| | 10 W to 1000 W (45 Hz to 1 kHz, Power factor = 0.5) | 0.41 % + 0.58R Where R is resolution of unit under calibration | |
| Power Factor Generate ^{3,5} (Lead and Lag) | 0.2 to 0.6 0.6 to 1 | 0.007 0.006 | Using Electrical Multifunction Calibrator by Direct Method |
| Oscilloscopes ⁵ Time Base | 100 ns to 20 ms | 0.0005 % | Using Electrical Multifunction Calibrator with Scope by Direct Method |
| Amplitude (Sine Wave) | 100 mVp-p to 30 Vp-p (1 kHz) | 0.079 % + 0.058 mV | |
| Bandwidth | 1 MHz to 300 MHz | 0.00019 % | |
| DC Voltage Measure ^{4,5} | 0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V | 0.0010 % + 0.36 μV 0.00092 % + 1.1 μV 0.00092 % + 10 μV 0.0012 % + 0.11 mV 0.0012 % + 1.1 mV | Using Digit Multimeter by Direct Method |
| DC Resistance Measure ^{4,5} | 0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ | 0.0017 % + 0.11 mΩ 0.0014 % + 1.2 mΩ 0.0012 % + 9.1 mΩ 0.0012 % + 52 mΩ 0.0012 % + 0.51 Ω 0.0018 % + 9.5 Ω 0.0058 % + 0.83 kΩ 0.058 % + 13 kΩ 1.6 % + 1.0 MΩ | Using Digit Multimeter by Direct Method |
| DC Current Measure ^{4,5} | 0 μA to 100 μA 100 μA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 1.9 A 1.9 A to 5 A 5 A to 10 A | 0.0023 % + 1.1 nA 0.0023 % + 7.6 nA 0.0023 % + 76 nA 0.0040 % + 0.77 μA 0.013 % + 17 μA 0.12 % + 0.70 mA 0.17 % + 3.1 mA 0.17 % + 3.1 mA | Using Digit Multimeter by Direct Method |
| | 10 A to 20 A 20 A to 50 A 50 A to 100 A | 5 mA 30 mA 60 mA | |
| | | | Using AC/DC Current Shunt & Digit Multimeter by Direct Method |

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| MEASURED QUANTITY or DEVICE TYPE CALIBRATED | RANGE | UNCERTAINTY ^{1,2} (±) | CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL) |
|--|---|---|--|
| DC Current Measure ^{4,5} (continued) | 100 A to 500 A 500 A to 1000 A | 12 A 24 A | Using Calibrated Clamp meter & Digit Multimeter by Direct Method |
| AC Voltage Measure ^{4,5} | (40 Hz to 1 kHz) 0 mV to 10 mV 10 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 700 V | 0.04 % + 1.5 µV 0.008 % + 5.2 µV 0.008 % + 38 µV 0.008 % + 0.56 mV 0.03 % + 4.3 mV 0.05 % + 39 mV | Using Digit Multimeter by Direct Method |
| AC Current Measure ^{4,5} | (10 Hz to 1 kHz) 10 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 1.9 A 1.9 A to 10 A | 0.07 % + 41 nA 0.04 % + 0.29 µA 0.04 % + 2.7 µA 0.04 % + 27 µA 0.12 % + 0.29 mA 0.17 % + 4.7 mA 0.17 % + 5.5 mA | Using Digit Multimeter by Direct Method |
| | (45 Hz to 500 Hz) 100 A to 500 A 500 A to 1000 A 1000 A to 2000 A | 19 A 47 A 79 A | Using Calibrated Clamp meter & Digit Multimeter by Direct Method |
| Capacitance Measure ^{4,5} @1 kHz | 1 nF to 10 nF 10 nF to 10 mF 10 mF to 100 mF | 1.6 % + 23 pF 0.8 % + 49 pF 3.1 % + 0.96 mF | Using Digit Multimeter by Direct Method |
| Time and Frequency | | | |
| Frequency Generate ^{3,5} | 10.0 Hz to 1199 kHz | 0.00019 % | Using Electrical Multifunction Calibrator by Direct Method |
| Frequency Measure ^{4,5} | 10 Hz to 10 MHz | 0.012 % | Using Digit Multimeter by Direct Method |
| Tachometer (Non-contact Type) ⁵ | 60 rpm to 90000 rpm | 0.00019 % | Using Electrical Multifunction Calibrator by Direct Method |

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

³Capability is suitable for the calibration of measuring devices in the stated ranges.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

⁴Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

⁵Also available as on-site calibration. Note that actual measurement uncertainties achievable at site outside the laboratory can normally be expected to be larger than the uncertainties listed on this Scope of Accreditation.

Notes:

CW = clockwise

CCW = counter-clockwise

p-p = peak to peak