



# CERTIFICATE OF ACCREDITATION

*This is to attest that*

**UNIVERSAL INSPECTION CO. LTD,**  
ROAD 3419, BLOCK 334, BUILDING 655, AL MAHUZ  
MANAMA, 334, KINGDOM OF BAHRAIN

**Calibration Laboratory CL-217**

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date February 1, 2023

Expiration Date February 1, 2025



A handwritten signature in black ink, reading 'Raj Nathan'.

**President**

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

## UNIVERSAL INSPECTION CO. LTD,

[www.ui.com.sa](http://www.ui.com.sa)

**Contact Name** Mr. Dinesh Kumar Kesavan

**Contact Phone** +966-508836773

*Accredited to ISO/IEC 17025:2017*

*Effective Date February 1, 2023*

### CALIBRATION AND MEASUREMENT CAPABILITY (CMC)\*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Dimensional</b>			
Vernier Caliper (Digital, Dial, Analog)	0 mm to 600 mm	0.040 mm	Using Caliper Checker & Length Bar as per JIS B 7507
Height Gauge (Digital/Analog)	0 mm to 600 mm	0.12 mm	Using Caliper Checker & Length Bar as per JIS B 7517
External Micrometer	0 mm to 25 mm 25 mm to 50 mm	0.002mm 0.004 mm	Using Gauge Blocks & Length Bars as per BS 870
Plunger Dial Gauge, Lever Dial Gauge (Digital/Analog)	0 mm to 25 mm	0.006 mm	Using Dial Gauge Calibrator as per JIS B 7503
<b>Mechanical</b>			
Universal Testing Machine (Compression only), Compression Testing Machine	0 kN to 300 kN 400 kN to 800 kN	0.95 kN 2.3 kN	Load Cell as per ISO 7500-1
Torque Wrench	0 N·m to 200 N·m 0 N·m to 1000 N·m	12 N·m 9.2 N·m	Torque Wrench Calibration System as per ISO 6789
Pressure Gauge / Pressure Transmitter (Hydraulic)	0 psi to 15000 psi 0 psi to 25000 psi	2.7 psi 3.8 psi	Dead Weight Tester High Pressure / Hydraulic Calibration Pump As per DKD R 6-1 & BS EN 837
Pressure Gauge - Pneumatic	0 bar to 50 bar	0.031 bar	High Pressure Pneumatic Calibration Pump as per BS EN 837
Vacuum Gauge	-0.95 bar to 0 bar	0.03 bar	By using Pressure Controller Procedure - ISO 3567
Sound Level Meter (1 kHz)	114 dB	0.99 dB	Sound Level Calibrator as per OIML R58

\* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Thermal</b>			
Infrared Thermometer	50 °C to 500 °C	1.6 °C	Using Black Body Furnace (e=0.98) by Direct Method as per ASTM E2847
Thermocouple Type K, Resistance Temperature Detector (RTD) Temperature Controller/Indicator with sensor, Temp. Recorder with sensor, Digital Data Logger with sensor, Thermometer with sensor, Temperature Transmitter	50 °C to 600 °C	1.0 °C	Using PRT with temp calibrator and Temperature Bath/ Dry block as per EURAMET-CG-11
Thermohygrometer	56 %RH to 80 %RH	1.8 %RH	Using Humidity Calibrator as per NIST SP-250-83
Temperature Bath, Oven, Furnace, Temperature Calibrator	50 °C to 600 °C	1.1 °C	Resistance Temperature Detector (RTD), S-Type Thermocouple and Temperature Calibrator, ASTM E145
Electrical Simulation of Thermocouples – Measure and Source Mode K-Type	-200 °C to 1300 °C	0.44 °C	Using Temperature Calibrator as per Euramet CG-11
<b>Electrical – DC/LF</b>			
DC Voltage Source <sup>3</sup>	0 µV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 240 V 240 V to 1000 V	0.023 mV 0.036 mV 0.000094 V 0.0013 V 0.0097 V 0.16 V	Clarke-Hess 8080 Multifunction Calibrator, Procedure UIC/P/EDMM based on Euramet CG-15
DC Voltage Measure <sup>4</sup>	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.0092 mV 0.00051 V 0.00089 V 0.0056 V 0.073 V	Precision Multimeter Fluke 8846A, Procedure UIC/P/EVARS based on Euramet CG-15
AC Voltage Source <sup>3</sup> (60 Hz)	1 mV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 240 V 240 V to 1000 V	0.081 mV 0.33 mV 0.0011 V 0.011 V 0.15 V 0.88 V	Clarke-Hess 8080 Multifunction Calibrator, Procedure UIC/P/EDMM based on Euramet CG-15
AC Voltage Measure <sup>4</sup> (60 Hz)	0 mV to 100 mV 100 mV to 1 V	0.031 mV 0.027 V	Precision Multimeter Fluke 8846A,

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
	1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.0025 V 0.026 V 0.38 V	Procedure UIC/P/EVARs based on Euramet CG-15
DC Current Source <sup>3</sup>	1 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	0.037 µA 0.00066 mA 0.0072 mA 0.036 mA 0.00050 A 0.0090 A	Clarke-Hess 8080 Multifunction Calibrator, Procedure UIC/P/EDMM based on Euramet CG-15
DC Current Measure <sup>4</sup>	0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 3 A 3 A to 10 A	0.096 µA 0.00067 mA 0.0082 mA 0.065 mA 0.46 mA 0.00086 A 0.0014 A 0.0064 A	Precision Multimeter Fluke 8846A, Procedure UIC/P/EVARs based on Euramet CG-15
AC Current Source <sup>3</sup> (60 Hz)	1 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	0.63 µA 0.0050 mA 0.023 mA 0.48 mA 0.011 A 0.052 A	Clarke-Hess 8080 Multifunction Calibrator, Procedure UIC/P/EDMM based on Euramet CG-15
AC Current Measure <sup>4</sup> (60 Hz)	0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 3 A 3 A to 10 A	0.42 µA 0.0015 mA 0.012 mA 0.099 mA 0.25 mA 0.0031 A 0.0044 A 0.010 A	Precision Multimeter Fluke 8846A, Procedure UIC/P/EVARs based on Euramet CG-15
DC Resistance Source <sup>3</sup>	0 kΩ to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 50 MΩ	0.00035 kΩ 0.0035 kΩ 0.035 kΩ 0.0010 MΩ 0.018 MΩ	Clarke-Hess 8080 Multifunction Calibrator, Procedure UIC/P/EDMM based on Euramet CG-15
DC Resistance Measure <sup>4</sup>	1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 50 MΩ	0.0018 kΩ 3.6 kΩ 3.6 kΩ 0.023 MΩ 0.30 MΩ	Precision Multimeter Fluke 8846A, Procedure UIC/P/EVARs 2-wire method based on Euramet CG-15
Holiday Detector	0 V to 30 kV	1.7 kV	AC/DC High Voltage Probe & High Voltage Divider, , ASTM D5162
Pin Hole Detector	0 V to 90 V	0.012 V	Precision Multimeter, NACE RP0188-99, ASTM D5162

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> ( $\pm$ )	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Electric & Diesel Welding Machine	Up to 600 A	3.9 A	ESAB Check Master 9000 as per BS EN 50504:2008
<i>Time and Frequency</i>			
Tachometer (Contact)	200 rpm to 1000 rpm 1000 rpm to 10000 rpm	1.5 rpm 2.7 rpm	Tachometer Calibrator as per SANAS TR-45-01

<sup>1</sup>The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

<sup>2</sup>When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

<sup>3</sup>Capability is suitable for the calibration of measuring devices in the stated ranges.

<sup>4</sup>Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.