



# CERTIFICATE OF ACCREDITATION

*This is to attest that*

## **UNIVERSAL INSPECTION CO.LTD.**

BLOCK NO.6, SHOP NO. A4, PLOT NO. 38  
EAST AHMADI 60006, STATE OF KUWAIT

### **Calibration Laboratory CL-222**

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date January 17, 2023

Expiration Date March 1, 2025



A handwritten signature in black ink, reading 'Raj Nathan'.

**President**

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

## UNIVERSAL INSPECTION CO.LTD.

[www.ui.com.sa](http://www.ui.com.sa)

**Contact Name** Mr. Dinesh Kumar Kesavan

**Contact Phone** +966-508836773r

Accredited to ISO/IEC 17025:2017

Effective Date January 17, 2023

### CALIBRATION AND MEASUREMENT CAPABILITY (CMC)\*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Dimensional</b>			
Vernier Caliper (Digital/Dial/Analog)	0 mm to 300 mm	73 µm	Using Caliper Checker & Length Bar as per JIS B 7507
Height Gauge (Digital/Analog)	0 mm to 300 mm	20 µm	Using Caliper Checker & Length Bar as per JIS B 7517
External Micrometer	0 mm to 25 mm 25 mm to 50 mm	2.6 µm 3.6 µm	Using Gauge Blocks & Length Bars as per BS 870
Dial Gauge (Digital/Analog)	0 mm to 25 mm	9.8 µm	Using Dial Gauge Calibrator as per JIS B 7503
<b>Mechanical</b>			
Torque Wrench	Up to 350 N·m 350 N·m to 900 N·m	5.8 N·m 36 N·m	Torque Wrench Calibration System as per ISO 6789
Tachometer	Contact – Up to 12000 rpm Photo – Up to 12000 rpm	1.3 rpm 1.8 rpm	Tachometer Calibrator as per ASTM F2046
Weighing Balance	1 mg to 500 mg 1 g to 200 g 1 kg 5 kg	0.09 mg 0.10 mg 0.12 g 0.49 g	Using F2 / M1 Class Weights as per OIML R 76
Pressure Gauge - Hydraulic	0 bar to 700 bar	6.0 bar	Reference Pressure gauge with Hydraulic Calibration Pump as per MSL Technical guide

\* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Pressure Gauge - Pneumatic	0 bar to 5 bar	0.08 bar	High Pressure Pneumatic Calibration Pump as per BS EN 837
Vacuum Gauge	0 bar to -0.95 bar	0.07 bar	Pressure Calibrator as per ISO 3567
Sound Level Meter(1 kHz)	94 dB 114 dB	1.1 dB 1.2 dB	Sound Level Calibrator as per ANSI S1.4
<b>Thermal</b>			
Thermocouple, Resistance Temperature Detector (RTD), Temperature Controller with Sensor, Temperature Indicator with Sensor, Temperature Recorder with Sensor, Digital Data Logger with Sensor, Thermometer with Sensor	-20 °C to 600 °C 600 °C to 1200 °C	0.89 °C 3.3 °C	Using Temperature Calibrator as per IEC 60751 & IEC 60584
Temperature bath, Oven, Furnace, Temperature Calibrator	-20 °C to 600 °C 600 °C to 1200 °C	0.73 °C 3.2 °C	Resistance Temperature Detector (RTD), S Type Thermocouple & Temperature Calibrator ASTM E145
Electrical Temperature Simulation – Measure <sup>4</sup> Thermocouples K- Type S- Type T-Type R-Type B-Type N-Type E-Type J-Type	-160 °C to 1200 °C 170 °C to 1750 °C -130 °C to 400 °C 150 °C to 1750 °C 950 °C to 1800 °C -100 °C to 950 °C -190 °C to 1200 °C	1.5 °C 2.3 °C 1.4 °C 2.3 °C 2.4 °C 1.2 °C 1.3 °C	Using Temperature Calibrator as per Euramet cg-11
Electrical Temperature Simulation – Generate <sup>3</sup> Thermocouples K- Type S- Type T-Type R-Type B-Type N-Type E-Type J-Type	-160 °C to 1200 °C 170 °C to 1750 °C -130 °C to 400 °C 150 °C to 1750 °C 950 °C to 1800 °C -100 °C to 950 °C -190 °C to 1200 °C	1.5 °C 2.4 °C 1.4 °C 2.4 °C 2.5 °C 1.1 °C 1.3 °C	

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Electrical – DC/LF</b>			
DC Voltage Source <sup>3</sup>	0 µV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 240 V 240 V to 1000 V	0.48 mV 2.3 mV 0.03 V 0.03 V 0.70 V 5.8 V	Multifunction Calibrator CLARKE- HESS 8080 Euramet cg-15
AC Voltage Source <sup>3</sup> (60 Hz)	0 mV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 240 V 240 V to 1000 V	12 µV 0.003 V 0.001 V 0.01 V 0.15 V 2.0 V	
DC Current Source <sup>3</sup>	1 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A	12 µA 12 mA 0.78 mA 0.01 A 0.02 A 0.04 A 0.06 A	
AC Current Source <sup>3</sup> (50 Hz)	1 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 10 A 10 A to 20 A	12 µA 0.17 mA 1.2 mA 12 mA 0.13 A 1.2 A 1.2 A	
DC Resistance Source <sup>3</sup>	1 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 50 MΩ 50 MΩ to 100 MΩ	0.04 Ω 0.02 kΩ 0.11 kΩ 0.15 kΩ 0.03 MΩ 0.58 MΩ 1.6 MΩ	
DC Voltage Measure <sup>4</sup>	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.31 mV 0.03 V 0.03 V 0.04 V 0.07 V	

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage Measure <sup>4</sup> (60 Hz)	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	53 µV 0.43 mV 0.02 V 0.43 V 4.6 V	Precision Multimeter Fluke 8846A as per Euramet cg-15
DC Current Measure <sup>4</sup>	0 µA to 100 µA 100 µA to 1 mA 0 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 3 A 3 A to 10 A	5.8 µA 10 µA 8.2 mA 0.02 mA 20 mA 0.03 A 0.04 A 0.06 A	
AC Current Measure <sup>4</sup> (60 Hz)	0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 3 A 3 A to 10 A	5.8 µA 0.16 mA 0.57 mA 12 mA 58 mA 0.06 A 0.57 A 1.2 A	
DC Resistance Measure <sup>4</sup>	1 Ω to 100 Ω 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 50 MΩ 50 MΩ to 100 MΩ	3.4 Ω 1.3 kΩ 14 kΩ 0.01 MΩ 0.03 MΩ 0.58 MΩ 1.2 MΩ	
DC Current Clamp Meter	20 A to 1000 A	4.2 A	
AC Current Clamp Meter	20 A to 1000 A	3.5 A	

<sup>1</sup>The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

<sup>2</sup>When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

<sup>3</sup>Capability is suitable for the calibration of measuring devices in the stated ranges.

<sup>4</sup>Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.