



# CERTIFICATE OF ACCREDITATION

*This is to attest that*

**UNIVERSAL INSPECTION CO.LTD.**

BLDG.NO.176, BLOCK NO.248, AZAIBA NORTH  
MUSCAT 130, SULTANATE OF OMAN

**Calibration Laboratory CL-221**

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date February 10, 2023

Expiration Date May 1, 2025



A handwritten signature in black ink, reading "Raj Nathan".

**President**

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

## UNIVERSAL TESTING CO.LTD.

[www.ui.com.sa](http://www.ui.com.sa)

**Contact Name** Dinesh Kumar Kesavan

**Contact Phone** +966-508836773

*Accredited to ISO/IEC 17025:2017*

*Effective Date February 10, 2023*

### CALIBRATION AND MEASUREMENT CAPABILITY (CMC)\*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> ( $\pm$ )	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Dimensional</b>			
Caliper (Digital, Dial, Vernier)	0 mm to 300 mm	17 $\mu$ m	Caliper Checker
Height Gauge (Digital, Dial, Vernier)	0 mm to 300 mm	13 $\mu$ m	Caliper Checker
Micrometer	2.5 mm to 25 mm	1.3 $\mu$ m	Gauge Block Set
Dial Gauge	0 mm to 25 mm	6.0 $\mu$ m	Dial Gauge Calibrator
<b>Mechanical</b>			
Pressure Gauge	Up to 1000 bar	1.5 bar	Dead Weight tester High Pressure Comparison Pump and Digital Pressure Gauge
Pressure Gauge Pneumatic	0 bar to 100 bar	0.02 bar	High Pressure Pneumatic Hand Pump and Digital Pressure Gauge
Vacuum Gauge	-0.85 bar to 0 bar	0.02 bar	High Pressure Pneumatic Hand Pump and Digital Vacuum Gauge
Weighing Balance	1 mg to 500 mg 1 g to 200 g 201 g to 5000 g	0.14 mg 160 mg 170 mg	E2 Class Weights
Tachometer (Contact Type)	12 rpm to 12000 rpm	1.0 rpm	Tachometer Calibrator
<b>Thermal</b>			
RTD / Thermocouple	-10 °C to 1200 °C	0.3 °C	Dry Bath Calibrator
Temperature Bath	50 °C to 600 °C	0.6 °C	SSPRT PT-100(4W) with Fluke DMM 8846A

\* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Electrical – DC/LF</b>			
DC Voltage – Source <sup>3</sup>	1 mV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 240 V 240 V to 1000 V	0.08 mV 0.60 mV 0.01 V 0.06 V 0.72 V 3.0 V	CLARK HESS Multifunction Electrical Calibrator
AC Voltage – Source <sup>3</sup> @50 Hz	1 mV to 20 mV 20 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 240 V 240 V to 1000 V	0.06 mV 0.60 mV 0.01 V 0.06 V 0.72 V 3.0 V	CLARK HESS Multifunction Electrical Calibrator
DC Current - Source <sup>3</sup>	1 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	0.60 µA 0.01 mA 0.06 mA 0.60 mA 0.01 A 0.06 A	CLARK HESS Multifunction Electrical Calibrator
AC Current – Source <sup>3</sup> @50 Hz	1 µA to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	0.62 µA 0.01 mA 0.06 mA 0.60 mA 0.01 A 0.06 A	CLARK HESS Multifunction Electrical Calibrator
Resistance - Source <sup>3</sup>	1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1000 MΩ	0.03 kΩ 0.30 kΩ 0.003 MΩ 0.03 MΩ 1.0 MΩ 23 MΩ	ZEAL ZMDRB Decade Resistance Box
DC Voltage -Measure <sup>4</sup>	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.30 mV 0.003 V 0.03 V 0.30 V 3.0 V	Fluke 8846A Precision Multimeter
AC Voltage- Measure <sup>4</sup> @50 Hz	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	0.29 mV 0.003 V 0.03 V 0.29 V 3.0 V	Fluke 8846A Precision Multimeter

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Current – Measure <sup>4</sup>	0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 3 A 3 A to 10 A	0.30 µA 0.003 mA 0.03 mA 0.30 mA 1.2 mA 0.003 A 0.009 A 0.03 A	Fluke 8846A Precision Multimeter
AC Current - Measure <sup>4</sup> @50Hz	0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 400 mA 400 mA to 1 A 1 A to 3 A 3 A to 10 A	0.30 µA 0.003 mA 0.03 mA 0.30 mA 1.2 mA 0.002 A 0.01 A 0.03 A	Fluke 8846A Precision Multimeter
DC Resistance – Measure <sup>4</sup>	1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1000 MΩ	0.03 kΩ 0.30 kΩ 0.003 MΩ 0.004 MΩ 0.30 MΩ 3.0 MΩ	Fluke 8846A Precision Multimeter
Electric & Diesel Welding Machine	Up to 600 A	0.84 A	ESAB Check Master 9000 as per BS EN 50504:2008

<sup>1</sup>The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

<sup>2</sup>When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

<sup>3</sup>Capability is suitable for the calibration of measuring devices in the stated ranges.

<sup>4</sup>Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.