



# CERTIFICATE OF ACCREDITATION

*This is to attest that*

## **CSA GROUP TESTING & CERTIFICATION INC. (POINTE-CLAIRE)**

865 ELLINGHAM AVENUE  
MONTREAL, QUEBEC H9R 5E8, CANADA

### **Calibration Laboratory CL-229**

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date December 4, 2023

Expiration Date August 1, 2025



A handwritten signature in black ink that reads "Raj Nathan".

**President**

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

## CSA GROUP TESTING & CERTIFICATION INC. (POINTE-CLAIRE)

[www.csagroup.org](http://www.csagroup.org)

**Contact Name** John Sena

**Contact Phone** +1-2165244990

*Accredited to ISO/IEC 17025:2017*

*Effective Date December 4, 2023*

### CALIBRATION AND MEASUREMENT CAPABILITY (CMC)\*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<i>Dimensional</i>			
Calipers	Up to 152.3 mm	6.9 µm	Direct measurement method using: Gauge Blocks and Caliper Checker
Outside Micrometers	Up to 25 mm	0.95 µm	Direct measurement method using: Gauge Blocks
Height Gauge	Up to 600 mm	15 µm	Direct measurement method using: Caliper Checker
Length – 1 Dimension	Up to 25 mm	0.002 mm	Direct measurement method using: Outside Micrometers
	Up to 150 mm	0.02 mm	Direct measurement method using: Calipers
	Up to 400 mm	0.02 mm	Direct measurement method using: Electronic Height Gages
Length – 3 Dimensions	Up to 605 mm	0.058 mm	Direct measurement method using: Portable CMM Faro Arm Edge
Levels and Protractors	0° to 90°	0.057°	Indirect Measurement Method with Gauge Blocks and Sine Bar

\* If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<b>Mechanical</b>			
Pressure Measuring Devices	-12 psi to 300 psi	0.17 psi	Direct measurement method using: Pressure Calibrator Fluke 718 300G
	300 psi to 5000 psi	0.080 %	
Scales & Balances	Up to 1000 g 1000 g to 3000 g 3000 g to 6000 g	0.06 g 0.07 g 0.12 g	Direct measurement method using: Reference Weights ASTM Class 1
	Up to 5 kg 5 kg to 20 kg 20 kg to 30 kg	0.58 g 0.61 g 0.65 g	
Weights	Up to 6000 g	0.2 g	Direct measurement method using: Ohaus SPX6201
	Up to 30 kg	2 g	
<b>Thermal</b>			
Temperature - Simulate Type B Thermocouple  Type C Thermocouple  Type E Thermocouple  Type J Thermocouple	600 °C to 800 °C	0.44 °C	Direct measurement method using: Multifunction Calibrator Fluke 5522A
	800 °C to 1000 °C	0.34 °C	
	1000 °C to 1550 °C	0.30 °C	
	1550 °C to 1820 °C	0.33 °C	
	0 °C to 150 °C	0.30 °C	
	150 °C to 650 °C	0.26 °C	
	650 °C to 1000 °C	0.31 °C	
	1000 °C to 1800 °C	0.50 °C	
	1800 °C to 2316 °C	0.84 °C	
	-250 °C to -100 °C	0.50 °C	
	-100 °C to -25 °C	0.16 °C	
	-25 °C to 350 °C	0.14 °C	
350 °C to 650 °C	0.16 °C		
650 °C to 1000 °C	0.21 °C		
-210 °C to -100 °C	0.27 °C		
-100 °C to 30 °C	0.16 °C		
30 °C to 150 °C	0.14 °C		

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Type J Thermocouple continued	150 °C to 760 °C 760 °C to 1200 °C	0.17 °C 0.23 °C	Direct measurement method using: Multifunction Calibrator Fluke 5522A
Type K Thermocouple	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1372 °C	0.33 °C 0.18 °C 0.16 °C 0.26 °C 0.40 °C	
Type L Thermocouple	-200 °C to -100 °C -100 °C to 800 °C 800 °C to 900 °C	0.37 °C 0.26 °C 0.17 °C	
Type N Thermocouple	-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C	0.40 °C 0.22 °C 0.19 °C 0.18 °C 0.27 °C	
Type R Thermocouple	0 °C to 250 °C 250 °C to 400 °C 400 °C to 1000 °C 1000 °C to 1767 °C	0.57 °C 0.35 °C 0.33 °C 0.40 °C	
Type S Thermocouple	0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1400 °C 1400 °C to 1767 °C	0.47 °C 0.36 °C 0.37 °C 0.46 °C	
Type T Thermocouple	-250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C	0.63 °C 0.24 °C 0.16 °C 0.14 °C	
Type U Thermocouple	-200 °C to 0 °C 0 °C to 600 °C	0.56 °C 0.27 °C	
RTD - Pt 385, 100 Ω	-200 °C to -80 °C -80 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C 300 °C to 400 °C 400 °C to 630 °C 630 °C to 800 °C	0.05 °C 0.05 °C 0.07 °C 0.09 °C 0.10 °C 0.12 °C 0.23 °C	
RTD - Pt 3926, 100 Ω	-200 °C to -80 °C -80 °C to 0 °C 0 °C to 100 °C 100 °C to 300 °C	0.05 °C 0.05 °C 0.07 °C 0.09 °C	

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
RTD - Pt 3926, 100 Ω continued	300 °C to 400 °C	0.10 °C	Direct measurement method using: Multifunction Calibrator Fluke 5522A
	400 °C to 630 °C	0.12 °C	
RTD - Pt 3916, 100 Ω	-200 °C to -190 °C	0.25 °C	
	-190 °C to -80 °C	0.04 °C	
	-80 °C to 0 °C	0.05 °C	
	0 °C to 100 °C	0.06 °C	
	100 °C to 260 °C	0.07 °C	
	260 °C to 300 °C	0.08 °C	
	300 °C to 400 °C	0.09 °C	
	400 °C to 600 °C	0.10 °C	
RTD - Pt 385, 200 Ω	600 °C to 630 °C	0.23 °C	
	-200 °C to -80 °C	0.04 °C	
	-80 °C to 0 °C	0.04 °C	
	0 °C to 100 °C	0.04 °C	
	100 °C to 260 °C	0.05 °C	
	260 °C to 300 °C	0.12 °C	
RTD - Pt 385, 500 Ω	300 °C to 400 °C	0.13 °C	
	400 °C to 600 °C	0.14 °C	
	600 °C to 630 °C	0.16 °C	
	-200 °C to -80 °C	0.04 °C	
	-80 °C to 0 °C	0.05 °C	
	0 °C to 100 °C	0.05 °C	
	100 °C to 260 °C	0.06 °C	
RTD - Pt 385, 1000 Ω	260 °C to 300 °C	0.08 °C	
	300 °C to 400 °C	0.08 °C	
	400 °C to 600 °C	0.09 °C	
	600 °C to 630 °C	0.11 °C	
	-200 °C to -80 °C	0.03 °C	
	-80 °C to 0 °C	0.03 °C	
	0 °C to 100 °C	0.04 °C	
RTD - PtNi 385, 120 Ω (Ni120)	100 °C to 260 °C	0.05 °C	
	260 °C to 300 °C	0.06 °C	
	300 °C to 400 °C	0.07 °C	
	400 °C to 600 °C	0.07 °C	
	600 °C to 630 °C	0.23 °C	
	100 °C to 260 °C	0.14 °C	
RTD - Cu 427, 10 Ω	-100 °C to 260 °C	0.3 °C	

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Temperature Measuring Devices	-25 °C to 25 °C 25 °C to 150 °C	0.024 °C 0.030 °C	Comparison method using: Fluke 1586A Super-DAQ, SPRT Fluke 5628 and temperature bath
Temperature/Refrigeration Chambers	-70 °C to 230 °C 50 °C to 230 °C	0.35 °C 0.35 °C	Direct measurement using Fluke 1586A Super-DAQ and SPRT Fluke 5628
Relative Humidity – Measure <sup>4</sup>	15 %RH to 30 %RH (5 °C to 40 °C)  30 %RH to 95 %RH (32 °C to 40 °C)	1.4 %RH  2.1 %RH	Direct measurement method using: Vaisala HMP76 and Vaisala MI70
<b>Electrical – DC/LF</b>			
DC Voltage – Measure <sup>4</sup>	10 mV to 100 mV 0.1 V to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	26 µV/V 5.2 µV/V 4.2 µV/V 6.3 µV/V 7.4 µV/V	Direct measurement method using: Reference DMM Fluke 8588A
DC High Voltage – Measure <sup>4</sup>	1 kV to 50 kV	0.58 %	Direct measurement method using: Reference DMM HP34401A and HV Divider ROSS VD-60
DC Voltage – Generate <sup>3</sup>	50 mV to 329.9999 mV 0.33 V to 3.299999 V 3.3 V to 32.99999 V 33 V to 329.9999 V 100 V to 1000.000 V	31 µV/V 12 µV/V 12 µV/V 16 µV/V 17 µV/V	Direct measurement method using: Multifunction Calibrator Fluke 5522A
	10 mV to 220 mV 0.22 V to 2.2 V 2.2 V to 11 V 11 V to 22 V 22 V to 220 V 220 V to 1100 V	69 µV/V 14 µV/V 7 µV/V 7 µV/V 9 µV/V 10 µV/V	Direct measurement method using: Multifunction Calibrator Fluke 5700A
DC Current – Measure <sup>4</sup>	1 µA to 10 µA 10 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 1 A to 10 A 10 A to 20 A  20 A to 54 A	730 µA/A 40 µA/A 40 µA/A 40 µA/A 89 µA/A 600 µA/A 860 µA/A  0.021 % + 1.2 mA	Direct measurement method using: Reference DMM Fluke 8588A  Indirect Measurements with Standard Resistor and DMM

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Current – Generate <sup>3</sup>	100 µA to 329.999 µA 0.33 mA to 3.299999 mA 3.3 mA to 32.9999 mA 33 mA to 329.9999 mA 0.33 A to 1.09999 A 1.1 A to 2.99999 A 3 A to 10.9999 A 11 A to 20.5 A	270 µA/A 160 µA/A 120 µA/A 120 µA/A 220 µA/A 320 µA/A 470 µA/A 820 µA/A	Direct measurement method using: Multifunction Calibrator Fluke 5522A <sup>4</sup>
	50 µA to 220 µA 0.22 mA to 2.2 mA 2.2 mA to 22 mA 22 mA to 220 mA 0.22 A to 2.2 A	200 µA/A 62 µA/A 62 µA/A 70 µA/A 120 µA/A	Direct measurement method using: Multifunction Calibrator Fluke 5700A
	20 A to 54 A	0.021 % + 1.2 mA	Indirect Measurements with Standard resistor, DMM and DC Power Supply
Clamp Meter Calibration	10 A to 16.4999 A 16.5 A to 149.999 A 150 A to 1025 A	0.5 % + 20 mA 0.5 % + 140 mA 0.5 % + 500 mA	Direct measurement method using: Multifunction Calibrator Fluke 5522A + Fluke 5500A/COIL 50-Turn
AC Voltage – Measure <sup>4</sup>	1 mV to 10 mV (20 Hz to 50 Hz) (60 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 1 MHz)	0.34 % 0.14 % 0.14 % 0.13 % 0.32 % 1.1 % 1.2 %	Direct measurement method using: Reference DMM Fluke 8588A
	10 mV to 100 mV (20 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 1 MHz)	760 µV/V 110 µV/V 200 µV/V 590 µV/V 0.27 % 1.2 %	
	0.1 V to 1 V (20 Hz to 2 kHz) (2 kHz to 10 kHz) (10 kHz to 30 kHz) (30 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 1 MHz)	130 µV/V 110 µV/V 0.23 % 0.53 % 0.59 % 1.2 %	

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage – Measure <sup>4</sup> continued	1 V to 10 V (20 Hz to 2 kHz)	100 µV/V	Direct measurement method using: Reference DMM Fluke 8588A
	(2 kHz to 10 kHz)	110 µV/V	
	(10 kHz to 30 kHz)	200 µV/V	
	(30 kHz to 100 kHz)	590 µV/V	
	(100 kHz to 1 MHz)	1.2 %	
10 V to 100 V (20 Hz to 2 kHz)	(2 kHz to 10 kHz)	100 µV/V	
	(10 kHz to 30 kHz)	94 µV/V	
	(30 kHz to 100 kHz)	200 µV/V	
		590 µV/V	
15 V (100 kHz to 1 MHz)		3.4 %	
22 V (100 kHz to 1 MHz)		2.6 %	
100 V to 1000 V (45 Hz to 2 kHz)		170 µV/V	
(2 kHz to 10 kHz)		170 µV/V	
AC High Voltage – Measure <sup>4</sup> (60 Hz)	1 kV to 42 kV	0.84 %	Direct measurement method using: Reference DMM HP34401A and HV Divider ROSS VD-60
AC Voltage – Generate <sup>3</sup> (Sine Wave)	1.0 mV to 32.999 mV (10 Hz to 45 Hz)	0.11 %	Direct measurement method using: Multifunction Calibrator Fluke 5522A Normal Output
	(45 Hz to 10 kHz)	580 µV/V	
	(10 kHz to 20 kHz)	620 µV/V	
	(20 kHz to 50 kHz)	0.12 %	
	(50 kHz to 100 kHz)	0.36 %	
	(100 kHz to 500 kHz)	1.0 %	
	33 mV to 329.999 mV (10 Hz to 45 Hz)	300 µV/V	
	(45 Hz to 10 kHz)	180 µV/V	
	(10 kHz to 20 kHz)	190 µV/V	
	(20 kHz to 50 kHz)	330 µV/V	
	(50 kHz to 100 kHz)	870 µV/V	
	(100 kHz to 500 kHz)	0.21 %	
	0.33 V to 3.29999 V (10 Hz to 45 Hz)	270 µV/V	
(45 Hz to 10 kHz)	160 µV/V		
(10 kHz to 20 kHz)	190 µV/V		
(20 kHz to 50 kHz)	270 µV/V		



# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
AC Voltage – Generate <sup>3</sup> (Sine Wave) (continued)	0.33 V to 3.29999 V (50 kHz to 100 kHz) (100 kHz to 500 kHz)	640 µV/V 0.23 %	Direct measurement method using: Multifunction Calibrator Fluke 5522A Normal Output	
	3.3 V to 32.9999 V (10 Hz to 45 Hz) (45 Hz to 10 kHz) (10 kHz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz)	280 µV/V 160 µV/V 230 µV/V 320 µV/V 820 µV/V		
	33 V to 329.999 V (45 Hz to 1 kHz) (1 kHz to 10 kHz) (10 kHz to 20 kHz)	160 µV/V 160 µV/V 240 µV/V		
	33 V to 329.999 V (20 kHz to 50 kHz) (50 kHz to 100 kHz)	280 µV/V 0.19 %		
	330 V to 1020.00 V (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	240 µV/V 240 µV/V 240 µV/V		
	22 mV to 220 mV (10 Hz to 20 Hz) (20 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 500 kHz) (500 kHz to 1 MHz)	0.17 % 540 µV/V 440 µV/V 630 µV/V 0.18 % 0.19 % 0.28 % 0.63 %		Direct Measurement with Multifunction Calibrator Fluke 5700A
	0.22 V to 2.2 V (10 Hz to 20 Hz) (20 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 500 kHz) (500 kHz to 1 MHz)	0.15 % 160 µV/V 71 µV/V 120 µV/V 280 µV/V 490 µV/V 0.12 % 0.26 %		
	2.2 V to 22 V (10 Hz to 20 Hz) (20 Hz to 40 Hz)	900 µV/V 160 µV/V		

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Voltage – Generate <sup>3</sup> (Sine Wave) (continued)	2.2 V to 22 V (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz) (300 kHz to 500 kHz) (500 kHz to 1 MHz)	71 µV/V 120 µV/V 250 µV/V 600 µV/V 0.15 % 0.30 %	Direct Measurement with Multifunction Calibrator Fluke 5700A
	22 V to 220 V (10 Hz to 20 Hz) (20 Hz to 40 Hz) (40 Hz to 20 kHz) (20 kHz to 50 kHz) (50 kHz to 100 kHz) (100 kHz to 300 kHz)	0.11 % 160 µV/V 78 µV/V 230 µV/V 540 µV/V 0.21 %	
	220 V to 1100 V (15 Hz to 50 Hz) (50 Hz to 1 kHz)	73 µV/V 73 µV/V	
AC Current – Measure <sup>4</sup>	30 µA to 100 µA (1 Hz to 2 kHz) (2 kHz to 10 kHz)	0.96 % 700 µA/A	Direct Measurement with Reference DMM Fluke 8588A
	100 µA to 1 mA (1 Hz to 2 kHz) (2 kHz to 10 kHz)	0.10 % 760 µA/A	
	1 mA to 10 mA (1 Hz to 2 kHz) (2 kHz to 10 kHz)	530 µA/A 760 µA/A	
	10 mA to 100 mA (1 Hz to 2 kHz) (2 kHz to 10 kHz)	510 µA/A 750 µA/A	
	100 mA to 1 A (1 Hz to 2 kHz) (2 kHz to 10 kHz)	760 µA/A 760 µA/A	
	1 A to 10 A (10 Hz to 2 kHz) (2 kHz to 10 kHz)	970 µA/A 970 µA/A	
	10 A to 20 A (10 Hz to 2 kHz) (2 kHz to 10 kHz)	0.18 % 0.22 %	
	20 A to 50 A (50/60 Hz)	0.16 %	
			Indirect Measurements with Standard Resistor and DMM

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Current – Generate <sup>3</sup>	29 µA to 329.99 µA		Direct Measurement with Multifunction Calibrator Fluke 5522A Inductive Compensation Off
	(10 Hz to 20 Hz)	0.25 %	
	(20 Hz to 45 Hz)	0.19 %	
	(45 Hz to 1 kHz)	0.17 %	
	(1 kHz to 5 kHz)	0.35 %	
	(5 kHz to 10 kHz)	0.78 %	
	(10 kHz to 30 kHz)	1.55 %	
	0.33 mA to 3.29999 mA		
	(10 Hz to 20 Hz)	0.22 %	
	(20 Hz to 45 Hz)	0.11 %	
	(45 Hz to 1 kHz)	0.09 %	
	(1 kHz to 5 kHz)	0.17 %	
	(5 kHz to 10 kHz)	0.41 %	
	(10 kHz to 30 kHz)	0.82 %	
	3.3 mA to 32.9999 mA		
	(10 Hz to 20 Hz)	0.20 %	
	(20 Hz to 45 Hz)	0.09 %	
	(45 Hz to 1 kHz)	0.05 %	
	(1 kHz to 5 kHz)	0.08 %	
	(5 kHz to 10 kHz)	0.18 %	
	(10 kHz to 30 kHz)	0.34 %	
	33 mA to 329.999 mA		
	(10 Hz to 20 Hz)	0.18 %	
	(20 Hz to 45 Hz)	0.09 %	
	(45 Hz to 1 kHz)	0.05 %	
	(1 kHz to 5 kHz)	0.12 %	
	(5 kHz to 10 kHz)	0.23 %	
	(10 kHz to 30 kHz)	0.47 %	
0.33 A to 1.09999 A			
(10 H to 45 Hz)	0.15 %		
(45 Hz to 1 kHz)	0.05 %		
(1 kHz to 5 kHz)	0.54 %		
(5 kHz to 10 kHz)	2.3 %		
1.1 A to 2.99999 A			
(10 Hz to 45 Hz)	0.14 %		
(45 Hz to 1 kHz)	0.05 %		
(1 kHz to 5 kHz)	0.50 %		
(5 kHz to 10 kHz)	2.1 %		
3 A to 10.9999 A			
(45 Hz to 100 Hz)	0.06 %		
(100 Hz to 1 kHz)	0.09 %		
(1 kHz to 5 kHz)	2.3 %		

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
AC Current – Generate <sup>3</sup> continued	11 A to 20.5 A (45 Hz to 100 Hz) (100 Hz to 1 kHz) (1 kHz to 5 kHz)	0.11 % 0.14 % 2.4 %	Direct Measurement with Multifunction Calibrator Fluke 5522A Inductive Compensation Of
	100 µA to 220 µA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	0.10 % 520 µA/A 280 µA/A 930 µA/A 0.22 %	Direct Measurement with Multifunction Calibrator Fluke 5700A
	0.22 mA to 22 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	0.11 % 360 µA/A 160 µA/A 930 µA/A 0.22 %	
	22 mA to 220 mA (10 Hz to 20 Hz) (20 Hz to 45 Hz) (45 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	0.15 % 360 µA/A 170 µA/A 930 µA/A 0.22 %	
	0.22 A to 2.2 A (20 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz)	610 µA/A 740 µA/A 0.79 %	
	20 A to 50 A (60 Hz/50 Hz)	0.21 %	Indirect Measurement with Standard Resistor, DMM and Rotek High Current Adapter
Clamp Meter Calibration (Torroidal and Non-Torroidal)	10 A to 16.499 A (45 Hz to 65 Hz) (65 Hz to 440 Hz)	0.56 % + 30 mA 1.0 % + 30 mA	Direct Measurement with Multifunction Calibrator Fluke 5522A + Fluke 5500A/COIL 50-Turn
	16.5 A to 149.999 A (45 Hz to 65 Hz) (65 Hz to 440 Hz)	0.56 % + 250 mA 0.79 % + 27 mA	
	150 A to 1025 A (45 Hz to 65 Hz) (65 Hz to 440 Hz)	0.56 % + 900 mA 1.0 % + 900 mA	
DC Resistance – Measure <sup>4</sup>	0.1 Ω to 1 Ω 1 Ω to 10 Ω	580 µΩ/Ω 120 µΩ/Ω	Direct Measurement with Reference DMM Fluke 8588A 4 W Ohms Function

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
DC Resistance – Measure <sup>4</sup> continued	10 Ω to 100 Ω	10 μΩ/Ω	Direct Measurement with Reference DMM Fluke 8588A 4 W Ohms Function
	100 Ω to 1 kΩ	9.9 μΩ/Ω	
	1 kΩ to 10 kΩ	9.6 μΩ/Ω	Direct Measurement with Reference DMM Fluke 8588A 2 Wire Ohms Function
	10 kΩ to 100 kΩ	10 μΩ/Ω	
	0.2 MΩ to 1 MΩ	30 μΩ/Ω	
	1 MΩ to 10 MΩ	180 μΩ/Ω	
	10 MΩ to 100 MΩ	0.18 %	
	100 MΩ to 1000 MΩ	0.63 %	
DC Power – Generate <sup>3</sup>	33 mV to 1020 V (0.33 mA to 329.99 mA) (0.33 A to 2.9999 A) (3 A to 20.5 A)	(relative to output in Watts) 0.023 % 0.022 % 0.07 %	Direct Measurement with Multifunction Calibrator Fluke 5522A
AC Power – Generate <sup>3</sup>	33 mV to 329.999 mV (3.3 mA to 8.999 mA) (9 mA to 32.999 mA)	(relative to output in Watts) 0.14 % 0.10 %	Direct Measurement with Multifunction Calibrator Fluke 5522A 45 Hz to 65 Hz, PF=1
	33 mV to 329.999 mV (33 mA to 89.99 mA) (90 mA to 329.99 mA)	0.14 % 0.10 %	
	330 mV to 1020 V (3.3 mA to 8.999 mA) (9 mA to 32.999 mA) (33 mA to 89.99 mA) (90 mA to 329.99 mA)	0.12 % 0.08 % 0.12 % 0.08 %	
	33 mV to 329.999 mV (0.33 A to 0.8999 A) (0.9 A to 2.1999 A) (2.2 A to 4.4999 A) (4.5 A to 20.5 A)	0.13 % 0.11 % 0.13 % 0.11 %	
	330 mV to 1020 V (0.33 A to 0.8999 A) (0.9 A to 2.1999 A) (2.2 A to 4.4999 A) (4.5 A to 20.5 A)	0.11 % 0.09 % 0.12 % 0.10 %	
Phase – Generate <sup>3</sup>	0° to 90° (10 Hz to 65 Hz) (65 Hz to 500 Hz) (500 Hz to 1 kHz) (1 kHz to 5 kHz) (5 kHz to 10 kHz) (10 kHz to 30 kHz)	0.1° 0.25° 0.5° 2.5° 5° 10°	Direct Measurement with Multifunction Calibrator Fluke 5522A

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)	
DC Resistance – Generate <sup>3</sup>	1 Ω to 10.9999 Ω	0.16 %	Direct Measurement with Multifunction Calibrator Fluke 5522A 4-WIRE Compensation 12 hrs zero cal	
	11 Ω to 32.9999 Ω	100 μΩ/Ω		
	33 Ω to 109.9999 Ω	43 μΩ/Ω		
	110 Ω to 329.9999 Ω	32 μΩ/Ω		
	330 Ω to 1.099999 kΩ	25 μΩ/Ω		
	1.1 kΩ to 3.299999 kΩ	32 μΩ/Ω		
	3.3 kΩ to 10.99999 kΩ	25 μΩ/Ω		
	11 kΩ to 32.99999 kΩ	32 μΩ/Ω		
	33 kΩ to 109.9999 kΩ	26 μΩ/Ω		
	110 kΩ to 329.99999 kΩ	35 μΩ/Ω		Direct Measurement with Multifunction Calibrator Fluke 5522A 2-WIRE Compensation 12 hrs zero cal
	330 kΩ to 1.099999 MΩ	30 μΩ/Ω		
	1.1 MΩ to 3.299999 MΩ	64 μΩ/Ω		
	3.3 MΩ to 10.99999 MΩ	110 μΩ/Ω		
	11 MΩ to 32.99999 MΩ	330 μΩ/Ω		
	33 MΩ to 109.9999 MΩ	430 μΩ/Ω		
110 MΩ to 329.9999 MΩ	0.29 %	Direct Measurement with Multifunction Calibrator Fluke 5700A 4-WIRE Compensation		
330 MΩ to 1100 MΩ	1.20 %			
1 Ω	85 μΩ/Ω			
1.9 Ω	85 μΩ/Ω			
10 Ω	26 μΩ/Ω			
19 Ω	24 μΩ/Ω			
100 Ω	16 μΩ/Ω			
190 Ω	16 μΩ/Ω			
1 kΩ	12 μΩ/Ω			
1.9 kΩ	12 μΩ/Ω			
10 kΩ	11 μΩ/Ω			
19 kΩ	11 μΩ/Ω			
100 kΩ	12 μΩ/Ω			
190 kΩ	14 μΩ/Ω			
1 MΩ	18 μΩ/Ω			
1.9 MΩ	26 μΩ/Ω			
10 MΩ	37 μΩ/Ω	Direct Measurement with Yokogawa 2792 Standard Resistors		
19 MΩ	100 μΩ/Ω			
100 MΩ	100 μΩ/Ω			
1 mΩ	0.02 %			
10 mΩ	0.10 %			
Capacitance – Generate <sup>3,5</sup>	1 Ω	0.05 %	Direct Measurement with Multifunction Calibrator Fluke 5522A <sup>3</sup> 2-WIRE Compensation 12 hrs zero cal	
	10 Ω	0.03 %		
	100 Ω	0.015 %		
	3.3 nF to 10.9999 nF	0.27 %		
	11 nF to 32.9999 nF	0.45 %		
	33 nF to 109.999 nF	0.27 %		
	110 nF to 329.999 nF	0.27 %		
	0.33 μF to 1.09999 μF	0.27 %		
1.1 μF to 3.29999 μF	0.27 %			
3.3 μF to 10.9999 μF	0.27 %			
11 μF to 32.9999 μF	0.39 %			

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Capacitance – Generate <sup>3,5</sup> continued	33 µF to 109.999 µF 110 µF to 329.999 µF 0.33 mF to 1.09999 mF 1.1 mF to 3.29999 mF 3.3 mF to 10.9999 mF 11 mF to 32.9999 mF 33 mF to 110 mF	0.43 % 0.43 % 0.43 % 0.43 % 0.43 % 0.66 % 0.93 %	Direct Measurement with Multifunction Calibrator Fluke 5522A <sup>3</sup> 2-WIRE Compensation 12 hrs zero cal
Impedance (Capacitance & Inductance) – Measure <sup>4</sup>	(20 Hz to 1 kHz) 100 Ω to 1 MΩ  (20 Hz to 10 kHz) 0.01 mΩ to 2.5 Ω 2.5 Ω to 100 Ω 1 MΩ to 4 MΩ 4 MΩ to 100 MΩ  (1 kHz to 10 kHz) 100 Ω to 1 MΩ  (10 kHz to 100 kHz) 0.01 mΩ to 2.5 Ω 2.5 Ω to 100 Ω 100 Ω to 25 kΩ 25 kΩ to 100 MΩ	0.05 % +  Z / 2 GΩ  0.3 % + 1 mΩ / Z  0.1 % + 1 mΩ / Z  0.1 % +  Z / 1.5 GΩ 0.2 % +  Z / 1.5 GΩ  0.1 % +  Z / 1.5 GΩ  0.5 % + 2 mΩ/ Z  0.2 % + 2 mΩ/ Z  0.2 % +  Z / 100 MΩ 0.5 % +  Z / 100 MΩ	Direct Measurement with LCR-Bridge R&S HM8118
Electric Energy	300 W·h	0.07 %	Direct Measurement with Fluke 5522A and “Internal Clock” from Yokogawa Digital Power Meter
Electric Charge	1 A·h	0.065 %	Direct Measurement with Fluke 5522A and “Internal Clock” from Yokogawa Digital Power Meter
<b>Time and Frequency</b>			
Frequency – Generate <sup>3</sup>	10 Hz 45 Hz 1 kHz 10 kHz 100 kHz 300 kHz 500 kHz	580 µHz/Hz 130 µHz/Hz 58 µHz/Hz 58 µHz/Hz 58 µHz/Hz 190 µHz/Hz 120 µHz/Hz	Direct Measurement with Multifunction Calibrator Fluke 5522A <sup>3</sup>
Frequency – Measure <sup>4</sup>	10 Hz 45 Hz 1 kHz 10 kHz 100 kHz	580 µHz/Hz 130 µHz/Hz 58 µHz/Hz 58 µHz/Hz 58 µHz/Hz	Direct Measurement with Reference DMM Fluke 8588A

# SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | [www.iasonline.org](http://www.iasonline.org)

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY <sup>1,2</sup> ( $\pm$ )	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
Frequency – Measure <sup>4</sup> continued	300 kHz 500 kHz 1 MHz	190 $\mu$ Hz/Hz 120 $\mu$ Hz/Hz 58 $\mu$ Hz/Hz	Direct Measurement with Reference DMM Fluke 8588A
Time – Measure <sup>4</sup>	100 s to 24 h	0.55 s	Direct Measurements NIST Special Publication 960-12
	5 s to 29 minutes	7.1 ms	Direct Measurement with Universal Counter Agilent 53131A and Fluke 5522A NIST Special Publication 960-12

<sup>1</sup>The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

<sup>2</sup>When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

<sup>3</sup>Capability is suitable for the calibration of measuring devices in the stated ranges.

<sup>4</sup>Capability is suitable for the calibration of devices intended to generate the indicated quantity in the stated ranges.

<sup>5</sup>The actual frequency applied by the calibrator cannot be selected and may be dependent on the measurement device under calibration. Approximate frequency ranges for a given capacitance or capacitance range may be found in the Fluke 552xA's published specifications.