



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

MERIT CALIBRATION

7923 WARNER AVENUE, SUITE K
HUNTINGTON BEACH, CALIFORNIA 92647, U.S.A.

Calibration Laboratory CL-234

has met the requirements of AC204, *IAS Accreditation Criteria for Calibration Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date April 6, 2023

Expiration Date December 1, 2024



A handwritten signature in black ink, reading "Raj Nathan".

President

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MERIT CALIBRATION

www.meritcalibration.com

Contact Name Brandon Howard

Contact Phone +1 657 259 0597

Accredited to ISO/IEC 17025:2017

Effective Date April 6, 2023

CALIBRATION AND MEASUREMENT CAPABILITY (CMC)*

MEASURED QUANTITY or DEVICE TYPE CALIBRATED	RANGE	UNCERTAINTY ^{1,2} (±)	CALIBRATION METHOD OR PROCEDURE, STANDARD EQUIPMENT (OPTIONAL)
<i>Thermal</i>			
Temperature Measuring Equipment –	32 °F to 212 °F	0.22 °F	Comparison Method MCP-1 RTD, Water Bath, Ice Point
Data Logger with External Probe & Sensor	-13 °F to 270 °F	0.19 °F	Comparison Method MCP-1 RTD, Drywell

¹The uncertainty covered by the Calibration and Measurement Capability (CMC) is expressed as the expanded uncertainty having a coverage probability of approximately 95 %. It is the smallest measurement uncertainty that a laboratory can achieve within its scope of accreditation when performing calibrations of a best existing device. The measurement uncertainty reported on a calibration certificate may be greater than that provided in the CMC due to the behavior of the calibration item and other factors that may contribute to the uncertainty of a specific calibration.

²When uncertainty is stated in relative terms (such as percent, a multiplier expressed as a decimal fraction or in scientific notation), it is in relation to instrument reading or instrument output, as appropriate, unless otherwise indicated.

** If information in this CMC is presented in non-SI units, the conversion factors stated in NIST Special Publication 811 "Guide for the Use of the International System of Units (SI)" apply.*