



INTERNATIONAL
ACCREDITATION
SERVICE®

CERTIFICATE OF ACCREDITATION

This is to attest that

MEASUREMENT CONTROL CENTER

187BD ACCRA ETD RD LOTIS
MOHAMMEDIA MAROC, 28810, MOROCCO

Testing Laboratory TL-1331

has met the requirements of AC89, *IAS Accreditation Criteria for Testing Laboratories*, and has demonstrated compliance with ISO/IEC Standard 17025:2017, *General requirements for the competence of testing and calibration laboratories*. This organization is accredited to provide the services specified in the scope of accreditation.

Effective Date May 28, 2025



International Accreditation Service
Issued under the authority of IAS management

Visit www.iasonline.org for current accreditation information.

SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

MEASUREMENT CONTROL CENTER

www.mccmaroc.com

Contact Name Hafid Mohamed

Contact Phone +212661081224

Accredited to ISO/IEC 17025:2017

Effective Date May 28, 2025

| | |
|--|--|
| Security Post Microbiological | |
| Number of particles depending on diameter | |
| NF EN ISO 14644-1 | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes. |
| NF EN ISO 14644-3 | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes. |
| Good Practices Manufacturing (BPF, B.O. 2016: L.D.1 - § 4, 5) | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes. |
| Downward speed | |
| NF EN 12469 | Measurement with a hot-wire anemometer |
| Good Practices Manufacturing (BPF, B.O. 2016: L.D.1 - § 3) | Measurement with a hot-wire anemometer |
| Volume flow | |
| NF EN 12469 | The airflow is measured using a hot-wire anemometer and multiplying the measurement by the cross-sectional area |
| Differential pressure | |
| NF EN 12469 | Measurement with pressure gauge |
| Filter integrity | |
| NF EN 12469 | Generation of an aerosol and measurement of penetration in downstream of the filter with a photometer |
| NF EN ISO 14644-3 | Generation of an aerosol and measurement of penetration in downstream of the filter with a photometer |



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| | |
|--|---|
| Clean rooms and environments controlled and related | |
| Number of particles depending on diameter | |
| NF EN ISO 14644-1 | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes |
| NF EN ISO 14644-3 | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes |
| Good Practices Manufacturing (BPF, B.O. 2016: L.D.1 - § 4, 5) | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes |
| Recovery time | |
| NF EN ISO 14644-3 | Particle counting at intervals of regular time after contamination using from an aerosol |
| Differential pressure | |
| NF EN ISO 14644-3 | Measuring with a pressure gauge |
| Temperature | |
| NF EN ISO 14644-3 | Measuring with a thermometer |
| Relative humidity | |
| NF EN ISO 14644-3 | Measure with a hygrometer |
| Air flow | |
| NF EN ISO 14644-3 | Measurement with a hot wire anemometer |
| Filter integrity | |
| NF EN ISO 14644-3 | Generation of an aerosol and measurement of penetration downstream of the filter with a photometer |
| Establishments health – Areas to environment mastered | |
| Number of particles depending on diameter | |
| NF S 90-351 | The method for calculating the particle count: A particle counter is used in order to obtain the concentration of airborne particles with a size equal to or greater than the specified sizes |
| Kinetics of decontamination | |
| NF S 90-351 | Particle counting at intervals of regular time after contamination using from an aerosol |



SCOPE OF ACCREDITATION

International Accreditation Service, Inc.

3060 Saturn Street, Suite 100, Brea, California 92821, U.S.A. | www.iasonline.org

| | |
|---|--|
| Air speed | |
| NF S 90-351 | The air velocity is measured with a hot-wire anemometer and a propeller anemometer. |
| Differential pressure | |
| NF S 90-351 | Measuring with a pressure gauge |
| Temperature | |
| NF S 90-351 | Measuring with a thermometer |
| Relative humidity | |
| NF S 90-351 | Measure with a hygrometer |
| Air flow | |
| NF S 90-351 | Measurement with a hot wire anemometer |
| Compressed air qualification | |
| Particle counting | |
| ISO 8573-1 | Measurement using a meter particle optics |
| Determination of oil content (aerosols and oil vapors) | |
| ISO 8573-2 | Measurement using fluorescence detectors |
| Determination of dew point/ humidity level | |
| ISO 8573-3 | Measurement using a dew point |
| Air & Surface Microbiological Sampling | |
| ISO 8573-7 | The method for verifying the presence of viable micro-organisms is to expose an agar nutrient to the compressed air sample |

